

INTOXILYZER 8000 INSPECTION

Serial No.: 80-002667

Location: Tox L

- | | Check When Done |
|---|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3455</u> Lot No. <u>12210</u> Rep AC <u>0.020</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3847</u> Lot No. <u>13020</u> Rep AC <u>0.300</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR3377</u> Lot No. <u>ICS#1</u> Rep AC <u>0.100AC + 0.05% ACETONE</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. <u>116110B0A2</u>	Cyl No. <u>133</u>	Exp. Date <u>06/01/2013</u>
Test 1 <u>0.080</u>	Test 4 <u>0.080</u>	Test 7 <u>0.080</u>
Test 2 <u>0.080</u>	Test 5 <u>0.080</u>	Test 8 <u>0.080</u>
Test 3 <u>0.079</u>	Test 6 <u>0.080</u>	Test 9 <u>0.080</u>
Average <u>0.080</u>		

L. Tank pressure, Level 3 Func D Sub G P: 370 psi Reg 375 psi

M. Remarks/Maintenance record: ANNUAL INSPECTION
REPLACED EXTERNAL BATTERY ON 10 MAY 13; REPLACED SIMULATOR
RETURN AND BREATH HOSE O-RING

13 MAY 13
Date

Charles E. Ed
Field Inspector's Signature

14 May 13
Date

Deb Kashner
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.13.00 06/09
05/13/2013 16:45

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 11611080A2
Standard Cyl #? 133
Standard Expiration? 06/01/2013
Oper No? 666666

Flow Cal. Date: 09/10/2007
Slope 693
Intercept -570251

IR Calibration Date: 06/03/2011
 3um 9um

0th Coef(*100): -3044 -42
1st Coef(*100): 248787 127659
2nd Coef(*100): 6023 1886
H2O adj(mg/l*10k): 632 468

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.13.00 06/09
05/13/2013 16:46

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:47
02 Std. Sol.	0.021	16:48
03 Room Air	0.000	16:49
04 Std. Sol.	0.022	16:49
05 Room Air	0.000	16:50
06 Std. Sol.	0.022	16:51
07 Room Air	0.000	16:51

08 Sim Temp = 34.0°C

Simul Ser No = DR3455

Std Sol No = 12210

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020AC (Sim. Sol.)

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.13.00 06/09
05/13/2013 16:52

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:53
02 Std. Sol.	0.302	16:53
03 Room Air	0.000	16:54
04 Std. Sol.	0.304	16:55
05 Room Air	0.000	16:55
06 Std. Sol.	0.305	16:56
07 Room Air	0.000	16:56

08 Sim Temp = 34.0°C

Simul Ser No = DR3847

Std Sol No = 13020

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *LINEARITY TEST*
0.300 AC (SIM. SOL.)

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.13.00 06/09
05/13/2013 16:57

Test	AC	Time
01 Room Air	0.000	16:58
02 *Subject Test	INT*	16:59
03 Room Air	0.000	16:59

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = QWERTYUIOPASDFG
Dr. Lic. = ND/DIS821456
Lot No = 11611080A2
Cyl No = 133
Expiration Date = 06/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

~~RFI CHECK~~ CEE
INTERFERENT CHECK

Form 106-I8000

0.100 AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.13.00 06/09
05/13/2013 16:59

Test	AC	Time
01 Diagnostic	OK	17:00
02 Room Air	RFI*	17:00
03 Room Air	0.000	17:01

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = MNBVCXZLKJH0987
Dr. Lic. = ND/DIS821456
Lot No = 11611080A2
Cyl No = 133
Expiration Date = 06/01/2013
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.13.00 06/09
05/13/2013 17:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:02
02 Std. Gas	0.080	17:02
03 Room Air	0.000	17:02
04 Std. Gas	0.080	17:03
05 Room Air	0.000	17:03
06 Std. Gas	0.079	17:04
07 Room Air	0.000	17:04

Lot No = 11611080A2
Cyl No = 133
Exp Date = 06/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080AC (E.G.S.)

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.13.00 06/09
05/13/2013 17:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:05
02 Std. Gas	0.080	17:05
03 Room Air	0.000	17:06
04 Std. Gas	0.080	17:06
05 Room Air	0.000	17:07
06 Std. Gas	0.080	17:07
07 Room Air	0.000	17:07

Lot No = 11611080A2
Cyl No = 133
Exp Date = 06/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080AC (E.G.S.)


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002667
Location = TOXL 8164.13.00 06/09
05/13/2013 17:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:08
02 Std. Gas	0.080	17:09
03 Room Air	0.000	17:09
04 Std. Gas	0.080	17:09
05 Room Air	0.000	17:10
06 Std. Gas	0.080	17:10
07 Room Air	0.000	17:11

Lot No = 11611080A2
Cyl No = 133
Exp Date = 06/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080AC (E.G.S.)

Form 106-I8000