

INTOXILYZER 8000 INSPECTION

Serial No.: 80-003065

Location: TOXL

- |  | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes?   | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?  | <u>✓</u>        |
| C. Is the breath tube heated?  | <u>✓</u>        |
| D. Is the diagnostic check complete?   | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?  | <u>✓</u>        |
| F. Print test. (Attach test record.)   | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5132</u> Lot No. <u>12210</u> Rep AC <u>0.020</u>  | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5143</u> Lot No. <u>13080</u> Rep AC <u>0.300</u>   | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR3846</u> Lot No. <u>    </u> Rep AC <u>0.100Ac + acetone</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)  | <u>✓</u>        |

K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.  
(Attach test record.)

Lot No. <u>23411080A1</u>	Cyl No. <u>15</u>	Exp. Date <u>10/01/2013</u>
Test 1 <u>0.081</u>	Test 4 <u>0.080</u>	Test 7 <u>0.081</u>
Test 2 <u>0.080</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.080</u>	Test 6 <u>0.080</u>	Test 9 <u>0.079</u>
Average <u>0.080</u>		

L. Tank pressure, Level 3 Func D Sub G P: 217 psi Reg 220 psi

M. Remarks/Maintenance record: Annual inspection, replaced external battery 13 May 13

DE-13  
Date 14 May 13

14 MAY 13  
Date

Deb Kashur  
Field Inspector's Signature  
Charles E. Ed  
Reviewed by



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003065  
Location = TOXL      8164.13.00 06/09  
05/14/2013      15:46

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:46
02 Std. Sol.	0.020	15:47
03 Room Air	0.000	15:48
04 Std. Sol.	0.020	15:48
05 Room Air	0.000	15:49
06 Std. Sol.	0.021	15:50
07 Room Air	0.000	15:50

08 Sim Temp = 34.0°C

Simul Ser No = DR5132  
Std Sol No = 12210  
County = 08      Oper No. = 777777

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:

Low AC      0.020AC

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003065  
Location = TOXL      8164.13.00 06/09  
05/14/2013      16:06

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:06
02 Std. Sol.	0.296	16:07
03 Room Air	0.000	16:07
04 Std. Sol.	0.296	16:08
05 Room Air	0.000	16:09
06 Std. Sol.	0.297	16:09
07 Room Air	0.000	16:10

08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 13080

County = 08

Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Linearity test 0.300 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003065  
Location = TOXL                8164.13.00 06/09  
05/14/2013                        16:17

Test	AC	Time
01 Room Air	0.000	16:18
02 *Subject Test	INT*	16:19
03 Room Air	0.000	16:19

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 15  
Expiration Date = 10/01/2013  
County = 08                          Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Interferent check 0.100 AC + acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-003065  
Location = TOXL                    8164.13.00 06/09  
05/14/2013                            16:19

Test	AC	Time
01 Diagnostic	OK	16:20
02 Room Air	RFI*	16:20
03 Room Air	0.000	16:21

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 23411080A1  
Cyl No = 15  
Expiration Date = 10/01/2013  
County = 08                            Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*RFI Test*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003065  
Location = TOXL      8164.13.00 06/09  
05/14/2013      16:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:22
02 Std. Gas	0.081	16:22
03 Room Air	0.000	16:22
04 Std. Gas	0.080	16:23
05 Room Air	0.000	16:23
06 Std. Gas	0.080	16:24
07 Room Air	0.000	16:24

Lot No = 23411080A1  
Cyl No = 15  
Exp Date = 10/01/2013  
County = 08      Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003065  
Location = TOXL      8164.13.00 06/09  
05/14/2013      16:25

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:25
02 Std. Gas	0.080	16:25
03 Room Air	0.000	16:26
04 Std. Gas	0.081	16:26
05 Room Air	0.000	16:27
06 Std. Gas	0.080	16:27
07 Room Air	0.000	16:28

Lot No = 23411080A1  
Cyl No = 15  
Exp Date = 10/01/2013  
County = 08      Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003065  
Location = TOXL      8164.13.00 06/09  
05/14/2013      16:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:30
02 Std. Gas	0.081	16:31
03 Room Air	0.000	16:31
04 Std. Gas	0.081	16:32
05 Room Air	0.000	16:32
06 Std. Gas	0.079	16:32
07 Room Air	0.000	16:33

Lot No = 23411080A1  
Cyl No = 15  
Exp Date = 10/01/2013  
County = 08      Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

Calibration check 0.080AC