



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (5-2011)

Chemical Test Operator Name (Print) <i>CHARLES E. EDER</i>		
Location <i>LISB (RAWSON COUNTY S.O.)</i>	Intoxilyzer® Serial Number <i>80-004205</i>	
Gas Lot Number <i>339130B0A4</i>	Gas Cylinder Number <i>021</i>	Gas Expiration Date <i>01/01/2016</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.081* AC
 - B. *0.081* AC
 - C. *0.081* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Breath Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Charles E. Eder</i>	Date <i>14 MAY 14</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Keshur</i>	Date <i>19 May 14</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = LISB 8164.13.00 06/09
05/14/2014 12:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:56
02 Std. Gas	0.081	12:56
03 Room Air	0.000	12:57
04 Std. Gas	0.081	12:57
05 Room Air	0.000	12:58
06 Std. Gas	0.081	12:58
07 Room Air	0.000	12:59

Lot No = 33913080A4
Cyl No = 21
Exp Date = 01/01/2016
County = 37 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CYLINDER INSTALL*
0.080 AC