



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59281 (03-2014)

Serial Number <u>80-005940</u>	Instrument Location <u>COOP</u>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. RFI Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if Any): _____
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <u>[Signature]</u>	Date <u>05/11/15 ✓</u>
Reviewed By (Crime Laboratory Use Only) <u>Deb Kasher</u>	Date <u>20 May 2015</u>

This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Certified By: [Signature] Certified Date: 15 Nov. 2016

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = COOP 8164.13.00 06/09
05/14/2015 19:45

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 13314080A2
Standard Cyl #? 60
Standard Expiration? 06/05/2016
Oper No? 100754

Flow Cal. Date: 06/21/2013
Slope 688
Intercept -659053

IR Calibration Date: 04/02/2015
 3um 9um

0th Coef(*100): -18917 -25088
1st Coef(*100): 267616 136158
2nd Coef(*100): 3374 1493
H2O adj(mg/l*10k): 427 417

***** Printer Test End *****



Operator Signature
DANIELLE M. ROSEWAREN

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = COOP 8164.13.00 06/09
05/14/2015 19:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:58
02 Std. Gas	0.078	19:59
03 Room Air	0.000	19:59
04 Std. Gas	0.078	20:00
05 Room Air	0.000	20:00
06 Std. Gas	0.078	20:00
07 Room Air	0.000	20:01

Lot No = 13314080A2
Cyl No = 60
Exp Date = 06/05/2016
County = 20 Oper No. = 100754



Operator Signature
DANIELLE M. ROSEWAREN

Remarks:

Intoxilyzer Test Record and Check
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005940
Location = COOP 8164.13.00 06/09
05/14/2015 20:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	20:02
02 Room Air	RFI*	20:02

*Invalid Test
Inhibited - RFI

Lot No = 13314080A2
Cyl No = 60
Exp Date = 06/05/2016
County = 20 Oper No. = 100754



Operator Signature
DANIELLE M. ROSEWAREN

Remarks: