



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59281 (5-2011)

Serial Number <i>80-004191</i>	Instrument Location <i>TRAIL Co. S.O.</i>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. RFI Test (Use CMS Mode; Key Radio During First Room Air).
- 7. Repair and/or Maintenance Performed (if Any): *CRIME LAB OR INSPECTION*
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>Ray Weber</i>	Date <i>05/15/2012</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashner</i>	Date <i>22 May 2012</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004191
Location = HILL 8164.13.00 06/09
05/15/2012 15:06

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 15410080A1
Standard Cyl #? 15
Standard Expiration? 07/01/2012
Oper No? 058405

Flow Cal. Date: 01/28/2010
Slope 738
Intercept -777574

IR Calibration Date: 01/28/2010
 3um 9um

0th Coef(*100): -23128 -33609
1st Coef(*100): 272485 131724
2nd Coef(*100): 1490 1056
H2O adj(mg/l*10k): 432 614

***** Printer Test End *****



Operator Signature
RAY G. WEBER

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004191
Location = HILL 8164.13.00 06/09
05/15/2012 15:26

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:27
02 Std. Gas	0.077	15:27
03 Room Air	0.000	15:28
04 Std. Gas	0.079	15:28
05 Room Air	0.000	15:29
06 Std. Gas	0.078	15:29
07 Room Air	0.000	15:30

Lot No = 15410080A1

Cyl No = 15

Exp Date = 07/01/2012

County = 49

Oper No. = 058405


Operator Signature
RAY G. WEBER

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004191
Location = HILL 8164.13.00 06/09
05/15/2012 15:54

Test	AC	Time
01 Diagnostic	OK	15:57
02 Room Air	RFI*	15:57
03 Room Air	0.000	15:58

*Invalid Test
 Inhibited - RFI

Sub Name = TEST, SUBJECT R
Sub DOB = 01/01/2012
Sub Sex = Male Weight = 123
Test = OTH Cit = 1234567
Dr. Lic. = ND/1234567
Lot No = 15410080A1
Cyl No = 15
Expiration Date = 07/01/2012
County = 49 Oper No. = 058405

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
RAY G. WEBER

Remarks: