

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004206

Location: Wahpeton PD

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5132 Lot No. 12010 Rep AC 0.020 ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3803 Lot No. 12050 Rep AC 0.300 ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3846 Lot No. — Rep AC 0.100 ✓
- J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
- K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 15410080 A1 Cyl No. 32 Exp. Date 07/01/2012
Test 1 0.080 Test 4 0.079 Test 7 0.080
Test 2 0.080 Test 5 0.080 Test 8 0.079
Test 3 0.080 Test 6 0.080 Test 9 0.080
Average 0.080

L. Tank pressure, Level 3 Func D Sub G P: 672 psi Reg 670 psi

M. Remarks/Maintenance record: Annual inspection

15 May 2012
Date

Deb Kashner
Field Inspector's Signature

17 MAY 12
Date

Clara E. Egan
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004206
Location = WAHP 8164.13.00 06/09
05/15/2012 19:43

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 15410080A1
Standard Cyl #? 32
Standard Expiration? 07/01/2012
Oper No? 777777

Flow Cal. Date: 06/10/2009
Slope 690
Intercept -645883

IR Calibration Date: 06/07/2010
 3um 9um

0th Coef(*100): -17579 -21810
1st Coef(*100): 254057 127149
2nd Coef(*100): 4591 1650
H2O adj(mg/l*10k): 756 795

***** Printer Test End *****

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Print test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004206
Location = WAHP 8164.13.00 06/09
05/15/2012 19:56

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:56
02 Std. Sol.	0.019	19:57
03 Room Air	0.000	19:58
04 Std. Sol.	0.020	19:58
05 Room Air	0.000	19:59
06 Std. Sol.	0.018	20:00
07 Room Air	0.000	20:00


08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 39

Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

low AC 0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004206
Location = WAHP 8164.13.00 06/09
05/15/2012 20:01

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:01
02 Std. Sol.	0.301	20:02
03 Room Air	0.000	20:03
04 Std. Sol.	0.302	20:03
05 Room Air	0.000	20:04
06 Std. Sol.	0.304	20:05
07 Room Air	0.000	20:05

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 39

Oper No. = 777777

Deb Keshur
Operator Signature
DEB KASHUR

Remarks:

Linearity test ~~0.100 AC~~ ^{DK} ~~done~~
0.300 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004206
Location = WAHP 8164.13.00 06/09
05/15/2012 20:06

Test	AC	Time
01 Room Air	0.000	20:07
02 *Subject Test	INT*	20:07
03 Room Air	0.000	20:08

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 15410080A1
Cyl No = 32
Expiration Date = 07/01/2012
County = 39 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.100AC + acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004206
Location = WAHP 8164.13.00 06/09
05/15/2012 20:08

Test	AC	Time
01 Diagnostic	OK	20:09
02 Room Air	RFI*	20:09
03 Room Air	0.000	20:10

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 15410080A1
Cyl No = 32
Expiration Date = 07/01/2012
County = 39 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004206
Location = WAHP 8164.13.00 06/09
05/15/2012 19:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:45
02 Std. Gas	0.080	19:46
03 Room Air	0.000	19:46
04 Std. Gas	0.080	19:47
05 Room Air	0.000	19:47
06 Std. Gas	0.080	19:47
07 Room Air	0.000	19:48

Lot No = 15410080A1

Cyl No = 32

Exp Date = 07/01/2012

County = 39

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004206
Location = WAHP 8164.13.00 06/09
05/15/2012 19:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:49
02 Std. Gas	0.079	19:49
03 Room Air	0.000	19:50
04 Std. Gas	0.080	19:50
05 Room Air	0.000	19:50
06 Std. Gas	0.080	19:51
07 Room Air	0.000	19:51

Lot No = 15410080A1
Cyl No = 32
Exp Date = 07/01/2012
County = 39 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004206
Location = WAHP 8164.13.00 06/09
05/15/2012 19:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:52
02 Std. Gas	0.080	19:52
03 Room Air	0.000	19:53
04 Std. Gas	0.079	19:53
05 Room Air	0.000	19:54
06 Std. Gas	0.080	19:54
07 Room Air	0.000	19:55

Lot No = 15410080A1
Cyl No = 32
Exp Date = 07/01/2012
County = 39 Oper No. = 777777

Deb Keshur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 ac