

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004936

Location: LaMoure Co SO

- Check When Done
- A. Is the warm-up time less than 20 minutes? ✓
  - B. Is a three-pronged grounded outlet used? ✓
  - C. Is the breath tube heated? ✓
  - D. Is the diagnostic check complete? ✓
  - E. Is the time, date, and year correct (re-set if necessary)? ✓
  - F. Print test. (Attach test record.) ✓
  - G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)  
Sim SN DR5132 Lot No. 12010 Rep AC 0.020 ✓
  - H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)  
Sim SN DR3803 Lot No. 12050 Rep AC 0.300 ✓
  - I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.  
(Attach test record.)  
Sim SN DR3846 Lot No. — Rep AC 0.100 ✓
  - J. RFI check. Run CMS mode. Key radio on first room air.  
(Attach test record.) ✓
  - K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.  
(Attach test record.)

Lot No. 05911080A1 Cyl No. 35 Exp. Date 02/01/2013

Test 1 <u>0.077</u>	Test 4 <u>0.077</u>	Test 7 <u>0.079</u>
Test 2 <u>0.078</u>	Test 5 <u>0.077</u>	Test 8 <u>0.078</u>
Test 3 <u>0.077</u>	Test 6 <u>0.078</u>	Test 9 <u>0.077</u>

Average 0.078

L. Tank pressure, Level 3 Func D Sub G P: 892 psi Reg 900 psi

M. Remarks/Maintenance record: Annual inspection

15 May 2012  
Date

Deb Kashur  
Field Inspector's Signature

17 MAY 12  
Date

Charles E. Egan  
Reviewed by

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004936  
Location = LAMO                    8164.13.00 06/09  
05/15/2012                            15:48

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            05911080A1  
Standard Cyl #?            35  
Standard Expiration?        02/01/2013  
Oper No?                    777777

Flow Cal. Date:            01/18/2010  
Slope                        732  
Intercept                  -767769

IR Calibration Date:        06/09/2011  
                                  3um                    9um

-----  
0th Coef(\*100):            -2249                    -2916  
1st Coef(\*100):            245643                   121920  
2nd Coef(\*100):            3450                     1789  
H2O adj(mg/l\*10k):        557                      621

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
DEB KASHUR

Remarks:



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004936  
Location = LAMO      8164.13.00 06/09  
05/15/2012      16:08

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:08
02 Std. Sol.	0.022	16:09
03 Room Air	0.000	16:10
04 Std. Sol.	0.021	16:11
05 Room Air	0.000	16:11
06 Std. Sol.	0.021	16:12
07 Room Air	0.000	16:12

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 23

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Low AC      0.020 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004936  
Location = LAMO      8164.13.00 06/09  
05/15/2012      16:13

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:14
02 Std. Sol.	0.306	16:14
03 Room Air	0.000	16:15
04 Std. Sol.	0.305	16:16
05 Room Air	0.000	16:16
06 Std. Sol.	0.305	16:17
07 Room Air	0.000	16:18

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 23

Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Linearity test 0.300 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004936  
Location = LAMO                8164.13.00 06/09  
05/15/2012                        16:04

Test	AC	Time
01 Room Air	0.000	16:05
02 *Subject Test	INT*	16:05
03 Room Air	0.000	16:05

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 05911080A1  
Cyl No = 35  
Expiration Date = 02/01/2013  
County = 23                          Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Interferent check 0.100 AC + acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004936  
Location = LAMO      8164.13.00 06/09  
05/15/2012      16:06

Test	AC	Time
01 Diagnostic	OK	16:06
02 Room Air	RFI*	16:07
03 Room Air	0.000	16:07

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 05911080A1  
Cyl No = 35  
Expiration Date = 02/01/2013  
County = 23      Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
DEB KASHUR

Remarks:



Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004936  
Location = LAMO      8164.13.00 06/09  
05/15/2012      15:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:50
02 Std. Gas	0.077	15:50
03 Room Air	0.000	15:51
04 Std. Gas	0.078	15:51
05 Room Air	0.000	15:52
06 Std. Gas	0.077	15:52
07 Room Air	0.000	15:53

Lot No = 05911080A1

Cyl No = 35

Exp Date = 02/01/2013

County = 23

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004936  
Location = LAMO      8164.13.00 06/09  
05/15/2012      15:53

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:54
02 Std. Gas	0.077	15:54
03 Room Air	0.000	15:54
04 Std. Gas	0.077	15:55
05 Room Air	0.000	15:55
06 Std. Gas	0.078	15:56
07 Room Air	0.000	15:56

Lot No = 05911080A1

Cyl No = 35

Exp Date = 02/01/2013

County = 23

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004936  
Location = LAMO                    8164.13.00 06/09  
05/15/2012                            15:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:57
02 Std. Gas	0.079	15:57
03 Room Air	0.000	15:58
04 Std. Gas	0.078	15:58
05 Room Air	0.000	15:59
06 Std. Gas	0.077	15:59
07 Room Air	0.000	16:00

Lot No = 05911080A1  
Cyl No = 35  
Exp Date = 02/01/2013  
County = 23                            Oper No. = 777777

  
\_\_\_\_\_  
Operator Signature  
DEB KASHUR

Remarks:  
*Calibration check 0.080 AC*