



OFFICE OF ATTORNEY GENERAL
 Crime Laboratory Division
 2641 East Main Avenue
 P.O. Box 937
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Darcy W Aberle

Location: Williston LEC

Intoxilyzer Serial No.: 80 - 004189

Lot Number: 659357

Cylinder Number: 45

Expiration Date: 09/03/11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.077 %

Test 2: 0.077 %

Test 3: 0.077 %

(ATTACH TEST RECORD)

Darcy W Aberle
 Chemical Test Operator Signature

Tested 5/16/10
 Date Cylinder Changed

Deb Kashur

18 May 2010

Form 105-G
 Toxicology Laboratory
 Crime Laboratory Division

80-004189
 P. 83/84 KA


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004189
Location = WILL 8164.13.00 06/09
05/16/2010 18:09

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:10
02 Std. Gas	0.077	18:10
03 Room Air	0.000	18:11
04 Std. Gas	0.077	18:11
05 Room Air	0.000	18:11
06 Std. Gas	0.077	18:12
07 Room Air	0.000	18:12

Lot No = 659357
Cyl No = 45
Exp Date = 09/03/2011
County = 53 Oper No. = 010101


Operator Signature
DARCY ABERLE

Remarks:

Form 106-I8000

80-004189
P. 07/34 FH