



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Dean Haaland</u>		
Location: <u>Cass County Jail</u>		
Intoxilyzer® Serial Number: <u>80-004205</u>		
Gas Lot Number: <u>659357</u>	Gas Cylinder Number: <u>40</u>	Expiration Date: <u>9-3-2011</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: 0. <u>081</u> % Test 2: 0. <u>081</u> % Test 3: 0. <u>081</u> %		
(ATTACH TEST RECORD)		

Chemical Test Operator Signature

5-16-10
 Date

Deb Rashner
 Reviewed By

18 May 2010
 Date

80-004205
 P. 69/65 KM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = CASS 8164.13.00 06/09
05/16/2010 23:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	23:02
02 Std. Gas	0.081	23:03
03 Room Air	0.000	23:03
04 Std. Gas	0.081	23:04
05 Room Air	0.000	23:04
06 Std. Gas	0.081	23:05
07 Room Air	0.000	23:05

Lot No = 659357
Cyl No = 40
Exp Date = 09/03/2011
County = 09 Oper No. = 109409




Operator Signature
DEAN J. HAALAND

Remarks:

Form 106-I8000

I hereby certify that the information in the
attached document is a true and correct copy
of the original document received by this office



Deputy Sheriff
Cass County Sheriff's Office

80-004205
P. 65/65 YH