

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004204

Location: Cass Co Jail

- Check When Done
- A. Is the warm-up time less than 20 minutes? ✓
 - B. Is a three-pronged grounded outlet used? ✓
 - C. Is the breath tube heated? ✓
 - D. Is the diagnostic check complete? ✓
 - E. Is the time, date, and year correct (re-set if necessary)? ✓
 - F. Print test. (Attach test record.) ✓
 - G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5132 Lot No. 12010 Rep AC 0.020 ✓
 - H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3803 Lot No. 12050 Rep AC 0.300 ✓
 - I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3846 Lot No. — Rep AC 0.100 ✓
 - J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
 - K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 23411080A1 Cyl No. 2 Exp. Date 10/01/2013

Test 1 <u>0.080</u>	Test 4 <u>0.079</u>	Test 7 <u>0.080</u>
Test 2 <u>0.080</u>	Test 5 <u>0.079</u>	Test 8 <u>0.079</u>
Test 3 <u>0.080</u>	Test 6 <u>0.079</u>	Test 9 <u>0.080</u>

Average 0.080

L. Tank pressure, Level 3 Func D Sub G P: 632 psi Reg 630 psi

M. Remarks/Maintenance record: Annual inspection

16 May 2012
Date

Dob Kashur
Field Inspector's Signature

17 MAY 12
Date

Charles E. ...
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = CASS 8164.13.00 06/09
05/16/2012 10:02

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 23411080A1
Standard Cyl #? 2
Standard Expiration? 10/01/2013
Oper No? 777777

Flow Cal. Date: 07/13/2010
Slope 750
Intercept -815939

IR Calibration Date: 07/14/2011
 3um 9um

0th Coef(*100): -9686 -4209
1st Coef(*100): 259274 128761
2nd Coef(*100): 3107 1691
H2O adj(mg/l*10k): 532 548

***** Printer Test End *****



Operator Signature
DEB KASHUR

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = CASS 8164.13.00 06/09
05/16/2012 10:20

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:20
02 Std. Sol.	0.020	10:21
03 Room Air	0.000	10:22
04 Std. Sol.	0.021	10:23
05 Room Air	0.000	10:23
06 Std. Sol.	0.020	10:24
07 Room Air	0.000	10:24

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 09

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

DK
Calibration *Low AC 0.020 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = CASS 8164.13.00 06/09
05/16/2012 10:30

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:31
02 Std. Sol.	0.288	10:32
03 Room Air	0.000	10:32
04 Std. Sol.	0.289	10:33
05 Room Air	0.000	10:33
06 Std. Sol.	0.292	10:34
07 Room Air	0.000	10:35

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 09

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = CASS 8164.13.00 06/09
05/16/2012 10:16

Test	AC	Time
01 Room Air	0.000	10:17
02 *Subject Test	INT*	10:17
03 Room Air	0.000	10:18

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 2
Expiration Date = 10/01/2013
County = 09 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.100 Ac + acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = CASS 8164.13.00 06/09
05/16/2012 10:13

Test	AC	Time
01 Diagnostic	OK	10:14
02 Room Air	RFI*	10:14
03 Room Air	0.000	10:15

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 2
Expiration Date = 10/01/2013
County = 09 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

DK
Juler RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = CASS 8164.13.00 06/09
05/16/2012 10:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:03
02 Std. Gas	0.080	10:04
03 Room Air	0.000	10:04
04 Std. Gas	0.080	10:05
05 Room Air	0.000	10:05
06 Std. Gas	0.080	10:05
07 Room Air	0.000	10:06

Lot No = 23411080A1
Cyl No = 2
Exp Date = 10/01/2013
County = 09 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = CASS 8164.13.00 06/09
05/16/2012 10:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:07
02 Std. Gas	0.079	10:07
03 Room Air	0.000	10:08
04 Std. Gas	0.079	10:08
05 Room Air	0.000	10:09
06 Std. Gas	0.079	10:09
07 Room Air	0.000	10:10

Lot No = 23411080A1
Cyl No = 2
Exp Date = 10/01/2013
County = 09 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = CASS 8164.13.00 06/09
05/16/2012 10:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:10
02 Std. Gas	0.080	10:11
03 Room Air	0.000	10:11
04 Std. Gas	0.079	10:12
05 Room Air	0.000	10:12
06 Std. Gas	0.080	10:12
07 Room Air	0.000	10:13

Lot No = 23411080A1
Cyl No = 2
Exp Date = 10/01/2013
County = 09 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.08070

Form 106-I8000