

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004205 Location: LISB (LISBON)

- | | Check When Done |
|---|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.) Sim SN <u>DR3455</u> Lot No. <u>12210</u> Rep AC <u>0.020</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.) Sim SN <u>DR3897</u> Lot No. <u>13020</u> Rep AC <u>0.300</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode. (Attach test record.) Sim SN <u>DR5191</u> Lot No. <u>ICS #1</u> Rep AC <u>0.100 AC + ACETONE 0.05%</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air. (Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode. (Attach test record.) | |

Lot No. 23411080 A1 Cyl No. 06 Exp. Date 10/01/2013
Test 1 0.081 Test 4 0.080 Test 7 0.080
Test 2 0.080 Test 5 0.080 Test 8 0.080
Test 3 0.080 Test 6 0.080 Test 9 0.080
Average 0.080

L. Tank pressure, Level 3 Func D Sub G P: 969 psi Reg 975 psi

M. Remarks/Maintenance record: ANNUAL INSPECTION

15 MAY 13
Date

Charles E. Ed...
Field Inspector's Signature

16 May 13
Date

Deb Kashur
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = LISB 8164.13.00 06/09
05/15/2013 10:03

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 23411080A1
Standard Cyl #? 06
Standard Expiration? 10/01/2013
Oper No? 666666

Flow Cal. Date: 06/04/2009
Slope 704
Intercept -610062

IR Calibration Date: 06/03/2009
 3um 9um

0th Coef(*100): -16944 -29673
1st Coef(*100): 265939 132317
2nd Coef(*100): 1696 900
H2O adj(mg/l*10k): 490 583

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks: *PRINT TEST*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = LISB 8164.13.00 06/09
05/15/2013 10:22

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 10:23 |
| 02 Std. Sol. | 0.017 | 10:23 |
| 03 Room Air | 0.000 | 10:24 |
| 04 Std. Sol. | 0.018 | 10:25 |
| 05 Room Air | 0.000 | 10:25 |
| 06 Std. Sol. | 0.018 | 10:26 |
| 07 Room Air | 0.000 | 10:26 |


08 Sim Temp = 34.0°C

Simul Ser No = DR3455

Std Sol No = 12210

County = 37

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC*
0.020 AC (Sim. Sol.)

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = LISB 8164.13.00 06/09
05/15/2013 10:27

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 10:28 |
| 02 Std. Sol. | 0.292 | 10:28 |
| 03 Room Air | 0.000 | 10:29 |
| 04 Std. Sol. | 0.295 | 10:30 |
| 05 Room Air | 0.000 | 10:30 |
| 06 Std. Sol. | 0.296 | 10:31 |
| 07 Room Air | 0.000 | 10:31 |

08 Sim Temp = 34.0°C

Simul Ser No = DR3847
Std Sol No = 13020
County = 37 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *LINEARITY TEST*
0.300 AC (Sim. Sol.)


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = LISB 8164.13.00 06/09
05/15/2013 10:18

| Test | AC | Time |
|------------------|-------|-------|
| 01 Room Air | 0.000 | 10:19 |
| 02 *Subject Test | INT* | 10:19 |
| 03 Room Air | 0.000 | 10:20 |

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = QWERTYUIOPASDFG
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 06
Expiration Date = 10/01/2013
County = 37 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT CHECK*
0.100AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = LISB 8164.13.00 06/09
05/15/2013 10:16

| Test | AC | Time |
|---------------|-------|-------|
| 01 Diagnostic | OK | 10:17 |
| 02 Room Air | RFI* | 10:17 |
| 03 Room Air | 0.000 | 10:17 |

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = MNBVCXZLKJH0987
Dr. Lic. = ND/DIS821456
Lot No = 23411080A1
Cyl No = 06
Expiration Date = 10/01/2013
County = 37 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI. CHECK


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = LISB 8164.13.00 06/09
05/15/2013 10:04

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 10:05 |
| 02 Std. Gas | 0.081 | 10:05 |
| 03 Room Air | 0.000 | 10:05 |
| 04 Std. Gas | 0.080 | 10:06 |
| 05 Room Air | 0.000 | 10:06 |
| 06 Std. Gas | 0.080 | 10:07 |
| 07 Room Air | 0.000 | 10:07 |

Lot No = 23411080A1
Cyl No = 06
Exp Date = 10/01/2013
County = 37 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC (E.G.S.)

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = LISB 8164.13.00 06/09
05/15/2013 10:08

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 10:08 |
| 02 Std. Gas | 0.080 | 10:09 |
| 03 Room Air | 0.000 | 10:09 |
| 04 Std. Gas | 0.080 | 10:10 |
| 05 Room Air | 0.000 | 10:10 |
| 06 Std. Gas | 0.080 | 10:11 |
| 07 Room Air | 0.000 | 10:11 |

Lot No = 23411080A1
Cyl No = 06
Exp Date = 10/01/2013
County = 37 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC (E.G.S.)

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = LISB 8164.13.00 06/09
05/15/2013 10:11

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 10:12 |
| 02 Std. Gas | 0.080 | 10:12 |
| 03 Room Air | 0.000 | 10:13 |
| 04 Std. Gas | 0.080 | 10:13 |
| 05 Room Air | 0.000 | 10:14 |
| 06 Std. Gas | 0.080 | 10:14 |
| 07 Room Air | 0.000 | 10:14 |

Lot No = 23411080A1

Cyl No = 06

Exp Date = 10/01/2013

County = 37

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080AC (E.G.S.)

Form 106-I8000