



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59281 (03-2014)

Serial Number 80-004943	Instrument Location McKenzie County Sheriff's Office, Watford City, ND
Reason for Install/Repair <input type="checkbox"/> Install After Receiving From Crime Laboratory <input checked="" type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000. *N/A*
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level 1, Function P).
  - B. ACA Test (Level 1, Function C).
  - C. RFI Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if Any): \_\_\_\_\_
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>Dwight B. New</i>	Date <i>05/16/2014</i> ✓
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashner</i>	Date <i>22 May 14</i>

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004943  
Location = WATF                      8164.13.00 06/09  
05/16/2014                                  16:54

\*\*\*\*\*Printer Test\*\*\*\*\*

abcdefghijklmnopqrstuvxyz1234567890-=-|  
ABCDEFGHIJKLMNRPQRSTUVWXYZ!@#%\$^&\*()\_+?

abcdefghijklmnopqrstuvxyz1234567890-=-|  
ABCDEFGHIJKLMNRPQRSTUVWXYZ!@#%\$^&\*()\_+?

Current Instrument Setup

Data Entry Mode:                    Enabled  
Start Test Sequence:                DABACABA  
Display Prelim Rslt?                Yes  
Display Third Digit?                Yes  
Inhib Printer(Y/N)?                No  
Display Volume?                      No  
Disable On Memfull?                No  
# of Print Copies?                  1  
Select Std (D/W/I)?                Dry  
Standard Value?                      0.080  
Standard Lot #?                      16512080A2  
Standard Cyl #?                      62  
Standard Expiration?                08/01/2014  
Oper No?                              030924  
  
Flow Cal. Date:                      01/18/2010  
Slope                                 693  
Intercept                            -665387

IR Calibration Date:                05/07/2013  
    3um                      9um

0th Coef(*100):	-10324	-14602
1st Coef(*100):	231940	125849
2nd Coef(*100):	6794	2184
H2O adj(mg/l*10k):	586	616

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
DUSTIN NEWMAN

Remarks: Good Print Test


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004943  
Location = WATF                    8164.13.00 06/09  
05/16/2014                            16:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:55
02 Std. Gas	0.079	16:55
03 Room Air	0.000	16:56
04 Std. Gas	0.080	16:56
05 Room Air	0.000	16:57
06 Std. Gas	0.079	16:57
07 Room Air	0.000	16:57

Lot No = 16512080A2  
Cyl No = 62  
Exp Date = 08/01/2014  
County = 27                            Oper No. = 030924



Operator Signature  
DUSTIN NEWMAN

Remarks: *Good ACA Test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004943  
Location = WATF                      8164.13.00 06/09  
05/16/2014                              16:58

Test	AC	Time
01 Room Air	RFI*	16:59
02 Room Air	RFI*	17:00

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, TEST T  
Sub DOB = 03/25/1969  
Sub Sex = Male                      Weight = 180  
Test = DUI                              Cit = 000000001  
Dr. Lic. = MT/03251969111  
Lot No = 16512080A2  
Cyl No = 62  
Expiration Date = 08/01/2014  
County = 27                              Oper No. = 030924



Operator Signature  
DUSTIN NEWMAN

Remarks: Good RFI Test