

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004953

Location: LINT (EMMONS COUNTY S.O.)

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5191 Lot No. 15160 Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3847 Lot No. 16060 Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR5190 Lot No. ICS Solution 0.10AC + 0.05% ACETONE ✓
 - H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. <u>19615080A5</u>	Cyl No. <u>005</u>	Exp. Date <u>9/5/2017</u>
Test 1 <u>0.080</u>	Test 4 <u>0.080</u>	Test 7 <u>0.080</u>
Test 2 <u>0.080</u>	Test 5 <u>0.080</u>	Test 8 <u>0.080</u>
Test 3 <u>0.079</u>	Test 6 <u>0.081</u>	Test 9 <u>0.080</u>
Average <u>0.080</u>		

J. Tank pressure, Level 3 Func D Sub G P: 1014 psi Reg ~1000 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION, NEW CYLINDER

5/13/2016
Date

Charles E. Ehr
Field Inspector's Signature

16 May 16 ✓
Date

Deb Shanaver
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004953
Location = LINT 8164.13.00 06/09
05/13/2016 09:54

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-==|
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-==|
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 19615080A5
Standard Cyl #? 5
Standard Expiration? 09/05/2017
Oper No? 666666

Flow Cal. Date: 01/19/2010
Slope 691
Intercept -628068

IR Calibration Date: 04/07/2015
 3um 9um

0th Coef(*100): -18160 -22754
1st Coef(*100): 254326 129076
2nd Coef(*100): 3670 1866
H2O adj(mg/l*10k): 414 530

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004953
Location = LINT 8164.13.00 06/09
05/13/2016 09:59

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:59
02 Std. Sol.	0.019	10:00
03 Room Air	0.000	10:01
04 Std. Sol.	0.018	10:01
05 Room Air	0.000	10:02
06 Std. Sol.	0.019	10:02
07 Room Air	0.000	10:03

08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 15160

County = 15

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

LOW AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004953
Location = LINT 8164.13.00 06/09
05/13/2016 10:05

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:06
02 Std. Sol.	0.308	10:06
03 Room Air	0.000	10:07
04 Std. Sol.	0.310	10:08
05 Room Air	0.000	10:08
06 Std. Sol.	0.309	10:09
07 Room Air	0.000	10:09

08 Sim Temp = 34.0°C

Simul Ser No = DR3847
Std Sol No = 16060
County = 15 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *HIGH AC*
 0.300 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004953
Location = LINT 8164.13.00 06/09
05/13/2016 10:12

Test	AC	Time
01 Room Air	0.000	10:13
02 *Subject Test	INT*	10:14
03 Room Air	0.000	10:14

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 19615080A5
Cyl No = 5
Expiration Date = 09/05/2017
County = 15 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.10 AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004953
Location = LINT 8164.13.00 06/09
05/13/2016 10:16

Test	AC	Time
01 Diagnostic	OK	10:16
02 Room Air	0.000	10:17
03 *Subject Test	RFI*	10:17
04 Room Air	0.000	10:18

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 19615080A5
Cyl No = 5
Expiration Date = 09/05/2017
County = 15 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004953
Location = LINT 8164.13.00 06/09
05/13/2016 10:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:18
02 Std. Gas	0.080	10:19
03 Room Air	0.000	10:19
04 Std. Gas	0.080	10:19
05 Room Air	0.000	10:20
06 Std. Gas	0.079	10:20
07 Room Air	0.000	10:21

Lot No = 19615080A5
Cyl No = 5
Exp Date = 09/05/2017
County = 15 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004953
Location = LINT 8164.13.00 06/09
05/13/2016 10:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:22
02 Std. Gas	0.080	10:22
03 Room Air	0.000	10:22
04 Std. Gas	0.080	10:23
05 Room Air	0.000	10:23
06 Std. Gas	0.081	10:24
07 Room Air	0.000	10:24

Lot No = 19615080A5
Cyl No = 5
Exp Date = 09/05/2017
County = 15 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004953
Location = LINT 8164.13.00 06/09
05/13/2016 10:24

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:25
02 Std. Gas	0.080	10:25
03 Room Air	0.000	10:26
04 Std. Gas	0.080	10:26
05 Room Air	0.000	10:27
06 Std. Gas	0.080	10:27
07 Room Air	0.000	10:28

Lot No = 19615080A5
Cyl No = 5
Exp Date = 09/05/2017
County = 15 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
0.080 AC