



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

| | |
|---|---|
| Serial Number 80-004939 | Instrument Location Dickinson LEC |
| Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____ | |

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): NONE
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

| | |
|-------------------------------|------------------------|
| Field Inspector Signature | Date 5/17/22 |
|-------------------------------|------------------------|

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

| | |
|--|---------------------------------------|
| Reviewed/Certified By Janelle Portscheller | Certified Date 03 June 2022 |
|--|---------------------------------------|

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = DKPD 8164.14.00 09/16
05/17/2022 19:43

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 02621080A1
Standard Cyl #? 032
Standard Expiration? 03/05/2023
Oper No? 130935

Flow Cal. Date: 01/18/2010
Slope 724
Intercept -779308

IR Calibration Date: 04/05/2017
 3um 9um

0th Coef(*100): -24406 -20805
1st Coef(*100): 261875 135290
2nd Coef(*100): 2828 1427
H2O adj(mg/l*10k): 563 393

***** Printer Test End *****



Operator Signature
KEVIN ELDRIDGE

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = DKPD 8164.14.00 09/16
05/17/2022 19:56

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 19:57 |
| 02 Std. Gas | 0.081 | 19:57 |
| 03 Room Air | 0.000 | 19:58 |
| 04 Std. Gas | 0.082 | 19:58 |
| 05 Room Air | 0.000 | 19:59 |
| 06 Std. Gas | 0.081 | 19:59 |
| 07 Room Air | 0.000 | 20:00 |

Lot No = 02621080A1
Cyl No = 032
Exp Date = 03/05/2023
County = 45 Oper No. = 130935


Operator Signature
KEVIN ELDRIDGE

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = DKPD 8164.14.00 09/16
05/17/2022 20:03

| Test | AC | Time |
|------------------|-------|-------|
| 01 Room Air | 0.000 | 20:04 |
| 02 *Subject Test | RFI* | 20:04 |
| 03 Room Air | 0.000 | 20:05 |

*Invalid Test
Inhibited - RFI

Sub Name = TEST, TEST NMN
Sub DOB = 01/01/1980
Sub Sex = Male Weight = 165
Test = DUI Cit = N/A
Dr. Lic. = ND/ABC809999
Lot No = 02621080A1
Cyl No = 032
Expiration Date = 03/05/2023
County = 45 Oper No. = 130935


Operator Signature
KEVIN ELDRIDGE

Remarks:

Form 106-I8000