

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-003067

Location: HILL (TRAILL COUNTY)

- | | Check When Done |
|--|-----------------|
| A. Is the breath tube heated? | <u>✓</u> |
| B. Is the diagnostic check complete? | <u>✓</u> |
| C. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| D. Print test. (Attach test record.) | <u>✓</u> |
| E. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5115</u> Lot No. <u>12111</u> Rep AC <u>0.020</u> | <u>✓</u> |
| F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5143</u> Lot No. <u>13080</u> Rep AC <u>0.300</u> | <u>✓</u> |
| G. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR3375</u> Lot No. <u>IC5#2</u> Rep AC <u>0.100 AC + 0.05% ACETONE</u> | <u>✓</u> |
| H. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. 16512080A2 Cyl No. 012 Exp. Date 08/01/14
Test 1 0.079 Test 4 0.080 Test 7 0.080
Test 2 0.080 Test 5 0.080 Test 8 0.080
Test 3 0.079 Test 6 0.079 Test 9 0.079
Average 0.080

J. Tank pressure, Level 3 Func D Sub G P: 983 psi Reg 1000 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

15 MAY 14
Date

19 May 14 ✓
Date

Charles E. Edler
Field Inspector's Signature

Deb Kashner
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = HILL 8164.13.00 06/09
05/15/2014 08:23

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 16512080A2
Standard Cyl #? 12
Standard Expiration? 08/01/2014
Oper No? 666666

Flow Cal. Date: 01/23/2013
Slope 664
Intercept -509770

IR Calibration Date: 01/23/2013
 3um 9um

0th Coef(*100): -25461 -19867
1st Coef(*100): 280844 142398
2nd Coef(*100): 1665 764
H2O adj(mg/l*10k): 384 269

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = HILL 8164.13.00 06/09
05/15/2014 08:41

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:42
02 Std. Sol.	0.016	08:43
03 Room Air	0.000	08:43
04 Std. Sol.	0.018	08:44
05 Room Air	0.000	08:44
06 Std. Sol.	0.017	08:45
07 Room Air	0.000	08:46


08 Sim Temp = 34.0°C

Simul Ser No = DR5115

Std Sol No = 12111

County = 49

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: 0.020 AC Low AC
~~CALIBRATION @ CEG~~

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = HILL 8164.13.00 06/09
05/15/2014 08:48

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:49
02 Std. Sol.	0.293	08:50
03 Room Air	0.000	08:50
04 Std. Sol.	0.298	08:51
05 Room Air	0.000	08:51
06 Std. Sol.	0.298	08:52
07 Room Air	0.000	08:53

08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 13080

County = 49

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

~~Calibration Check CEE~~
~~CEE 0.000 AC~~ HIGH AC
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = HILL 8164.13.00 06/09
05/15/2014 08:27

Test	AC	Time
01 Room Air	0.000	08:28
02 *Subject Test	INT*	08:29
03 Room Air	0.000	08:29

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 16512080A2
Cyl No = 12
Expiration Date = 08/01/2014
County = 49 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

INTERFERENT CHECK

0.100AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = HILL 8164.13.00 06/09
05/15/2014 08:24

Test	AC	Time
01 Diagnostic	OK	08:26
02 Room Air	RFI*	08:26
03 Room Air	0.000	08:26

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 16512080A2
Cyl No = 12
Expiration Date = 08/01/2014
County = 49 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = HILL 8164.13.00 06/09
05/15/2014 08:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:32
02 Std. Gas	0.079	08:32
03 Room Air	0.000	08:32
04 Std. Gas	0.080	08:33
05 Room Air	0.000	08:33
06 Std. Gas	0.079	08:34
07 Room Air	0.000	08:34

Lot No = 16512080A2
Cyl No = 12
Exp Date = 08/01/2014
County = 49 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = HILL 8164.13.00 06/09
05/15/2014 08:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:37
02 Std. Gas	0.080	08:37
03 Room Air	0.000	08:37
04 Std. Gas	0.080	08:38
05 Room Air	0.000	08:38
06 Std. Gas	0.079	08:39
07 Room Air	0.000	08:39

Lot No = 16512080A2
Cyl No = 12
Exp Date = 08/01/2014
County = 49 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003067
Location = HILL 8164.13.00 06/09
05/15/2014 08:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:55
02 Std. Gas	0.080	08:55
03 Room Air	0.000	08:56
04 Std. Gas	0.080	08:56
05 Room Air	0.000	08:57
06 Std. Gas	0.079	08:57
07 Room Air	0.000	08:57

Lot No = 16512080A2
Cyl No = 12
Exp Date = 08/01/2014
County = 49 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC