

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-003072

Location: GARR

- |  | Check When Done |
|--|-----------------|
| A. Is the breath tube heated?  | <u>✓</u>        |
| B. Is the diagnostic check complete?   | <u>✓</u>        |
| C. Is the time, date, and year correct (re-set if necessary)?  | <u>✓</u>        |
| D. Print test. (Attach test record.)   | <u>✓</u>        |
| E. Low AC. Use $<0.03$ AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5115</u> Lot No. <u>12111</u> Rep AC <u>0.020</u>  | <u>✓</u>        |
| F. High AC. Use $\geq 0.25$ AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5143</u> Lot No. <u>13080</u> Rep AC <u>0.300</u>   | <u>✓</u>        |
| G. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR3315</u> Lot No. <u>1052</u> Rep AC <u>0.100AC + 0.05% acetone</u> | <u>✓</u>        |
| H. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)  | <u>✓</u>        |
| I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)  |                 |

Lot No. 16512080A2 Cyl No. 72 Exp. Date 08/01/2014

Test 1 <u>0.080</u>	Test 4 <u>0.080</u>	Test 7 <u>0.081</u>
Test 2 <u>0.079</u>	Test 5 <u>0.081</u>	Test 8 <u>0.080</u>
Test 3 <u>0.080</u>	Test 6 <u>0.081</u>	Test 9 <u>0.081</u>

Average 0.080

J. Tank pressure, Level 3 Func D Sub G P: 912 psi Reg 900 psi

K. Remarks/Maintenance record: Annual inspection

16 May 14  
Date

19 May 14  
Date

Deb Kaslow  
Field Inspector's Signature

Charles E. Ede  
Reviewed by

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-003072  
Location = GARR                    8164.13.00 06/09  
05/16/2014                            14:59

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?         1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            16512080A2  
Standard Cyl #?            72  
Standard Expiration?       08/01/2014  
Oper No?                    777777

Flow Cal. Date:            03/09/2011  
Slope                       682  
Intercept                  -620512

IR Calibration Date:        03/09/2011

	3um	9um
0th Coef(*100):	-24766	-17499
1st Coef(*100):	264559	138630
2nd Coef(*100):	1876	1114
H2O adj(mg/l*10k):	497	331

\*\*\*\*\* Printer Test End \*\*\*\*\*

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

Print Test

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003072  
Location = GARR      8164.13.00 06/09  
05/16/2014      15:14

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:15
02 Std. Sol.	0.018	15:15
03 Room Air	0.000	15:16
04 Std. Sol.	0.020	15:17
05 Room Air	0.000	15:17
06 Std. Sol.	0.018	15:18
07 Room Air	0.000	15:18


08 Sim Temp = 34.0°C

Simul Ser No = DR5115

Std Sol No = 12111

County = 28

Oper No. = 777777

  
\_\_\_\_\_  
Operator Signature  
DEB KASHUR

Remarks:

Low AC      0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003072  
Location = GARR      8164.13.00 06/09  
05/16/2014      15:20

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:21
02 Std. Sol.	0.299	15:22
03 Room Air	0.000	15:22
04 Std. Sol.	0.303	15:23
05 Room Air	0.000	15:23
06 Std. Sol.	0.304	15:24
07 Room Air	0.000	15:25

08 Sim Temp = 34.0°C

Simul Ser No = DR5143  
Std Sol No = 13080  
County = 28      Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

High AC 0.300AC

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003072  
Location = GARR                8164.13.00 06/09  
05/16/2014                        15:26

Test	AC	Time
01 Room Air	0.000	15:27
02 *Subject Test	INT*	15:27
03 Room Air	0.000	15:28

\*Invalid Test  
Interferent Detected

Sub Name = INTERF, NA NA  
Sub DOB = 08/01/1995  
Sub Sex = Unknown                Weight = NA  
Test = DUI                         Cit = NA  
Dr. Lic. = ND/NA  
Lot No = 16512080A2  
Cyl No = 72  
Expiration Date = 08/01/2014  
County = 28                        Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Interferent check 0.100AE + 0.05% acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003072  
Location = GARR                8164.13.00 06/09  
05/16/2014                      15:28

Test	AC	Time
01 Diagnostic	OK	15:29
02 Room Air	RFI*	15:29
03 Room Air	0.000	15:30

\*Invalid Test  
Inhibited - RFI

Sub Name = RFI, NA NA  
Sub DOB = 08/01/1975  
Sub Sex = Unknown                Weight = NA  
Test = DUI                         Cit = NA  
Dr. Lic. = ND/NA  
Lot No = 16512080A2  
Cyl No = 72  
Expiration Date = 08/01/2014  
County = 28                        Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
DEB KASHUR

Remarks:

*RFI check*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003072  
Location = GARR      8164.13.00 06/09  
05/16/2014      15:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:00
02 Std. Gas	0.080	15:01
03 Room Air	0.000	15:01
04 Std. Gas	0.079	15:02
05 Room Air	0.000	15:02
06 Std. Gas	0.080	15:02
07 Room Air	0.000	15:03

Lot No = 16512080A2  
Cyl No = 72  
Exp Date = 08/01/2014  
County = 28      Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:  
*Calibration check 0.080 AC*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003072  
Location = GARR      8164.13.00 06/09  
05/16/2014      15:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:04
02 Std. Gas	0.080	15:04
03 Room Air	0.000	15:05
04 Std. Gas	0.081	15:05
05 Room Air	0.000	15:06
06 Std. Gas	0.081	15:06
07 Room Air	0.000	15:06

Lot No = 16512080A2  
Cyl No = 72  
Exp Date = 08/01/2014  
County = 28      Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

Calibration check 0.080AC



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003072  
Location = GARR      8164.13.00 06/09  
05/16/2014      15:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:08
02 Std. Gas	0.081	15:08
03 Room Air	0.000	15:09
04 Std. Gas	0.080	15:09
05 Room Air	0.000	15:10
06 Std. Gas	0.081	15:10
07 Room Air	0.000	15:10

Lot No = 16512080A2  
Cyl No = 72  
Exp Date = 08/01/2014  
County = 28      Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check      0.080 AC*

Form 106-I8000