

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004205

Location: LISB (Ransom County)

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5115 Lot No. 12111 Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5143 Lot No. 13080 Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3375 Lot No. DR5112 Rep AC 0.100AC + 0.05% ACETONE ✓
 - H. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 33913080A4 Cyl No. 021 Exp. Date 01/01/2016
Test 1 0.080 Test 4 0.081 Test 7 0.081
Test 2 0.080 Test 5 0.080 Test 8 0.081
Test 3 0.081 Test 6 0.080 Test 9 0.081
Average 0.081

J. Tank pressure, Level 3 Func D Sub G P: 997 psi Reg 1000 psi

K. Remarks/Maintenance record: REPLACED BREATH HOSE O-RING, ANNUAL

INSPECTION

14 MAY 14
Date

19 May 14 ✓
Date

Charles E. Edin
Field Inspector's Signature

Deb Kashner
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = LISB 8164.13.00 06/09
05/14/2014 13:01

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 33913080A4
Standard Cyl #? 21
Standard Expiration? 01/01/2016
Oper No? 666666

Flow Cal. Date: 06/04/2009
Slope 704
Intercept -610062

IR Calibration Date: 06/03/2009
 3um 9um

0th Coef(*100): -16944 -29673
1st Coef(*100): 265939 132317
2nd Coef(*100): 1696 900
H2O adj(mg/l*10k): 490 583

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = LISB 8164.13.00 06/09
05/14/2014 13:08

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:09
02 Std. Sol.	0.018	13:10
03 Room Air	0.000	13:10
04 Std. Sol.	0.019	13:11
05 Room Air	0.000	13:11
06 Std. Sol.	0.019	13:12
07 Room Air	0.000	13:12
08 Sim Temp = 34.0°C		

Simul Ser No = DR5115
Std Sol No = 12111
County = 37 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *LOW AC*
0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = LISB 8164.13.00 06/09
05/14/2014 13:18

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:19
02 Std. Sol.	0.291	13:20
03 Room Air	0.000	13:20
04 Std. Sol.	0.291	13:21
05 Room Air	0.000	13:22
06 Std. Sol.	0.290	13:22
07 Room Air	0.000	13:23

08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 13080

County = 37

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *HIGH AC*
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = LISB 8164.13.00 06/09
05/14/2014 13:04

Test	AC	Time
01 Room Air	0.000	13:05
02 *Subject Test	INT*	13:06
03 Room Air	0.000	13:06

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 33913080A4
Cyl No = 21
Expiration Date = 01/01/2016
County = 37 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT CHECK*
CEE
0.080 AC + 0.05% ACETONE
0.100
Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = LISB 8164.13.00 06/09
05/14/2014 13:01

Test	AC	Time
01 Diagnostic	OK	13:02
02 Room Air	RFI*	13:03
03 Room Air	0.000	13:03

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 33913080A4
Cyl No = 21
Expiration Date = 01/01/2016
County = 37 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = LISB 8164.13.00 06/09
05/14/2014 13:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:31
02 Std. Gas	0.080	13:31
03 Room Air	0.000	13:32
04 Std. Gas	0.080	13:32
05 Room Air	0.000	13:32
06 Std. Gas	0.081	13:33
07 Room Air	0.000	13:33

Lot No = 33913080A4
Cyl No = 21
Exp Date = 01/01/2016
County = 37 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = LISB 8164.13.00 06/09
05/14/2014 13:34

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:34
02 Std. Gas	0.081	13:35
03 Room Air	0.000	13:35
04 Std. Gas	0.080	13:35
05 Room Air	0.000	13:36
06 Std. Gas	0.080	13:36
07 Room Air	0.000	13:37

Lot No = 33913080A4
Cyl No = 21
Exp Date = 01/01/2016
County = 37 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = LISB 8164.13.00 06/09
05/14/2014 13:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:38
02 Std. Gas	0.081	13:38
03 Room Air	0.000	13:39
04 Std. Gas	0.081	13:39
05 Room Air	0.000	13:39
06 Std. Gas	0.081	13:40
07 Room Air	0.000	13:40

Lot No = 33913080A4
Cyl No = 21
Exp Date = 01/01/2016
County = 37 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000