

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004944

Location: WFPD (WEST FARGO P.D.)

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5115 Lot No. 12111 Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5143 Lot No. 13080 Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3375 Lot No. ICS #2 Rep AC 0.100 AC + 0.05% ACETONE ✓
 - H. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 16512080A2 Cyl No. 018 Exp. Date 08/01/2014

Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.082</u>	Test 8 <u>0.083</u>
Test 3 <u>0.081</u>	Test 6 <u>0.081</u>	Test 9 <u>0.081</u>

Average 0.081

J. Tank pressure, Level 3 Func D Sub G P: 878 psi Reg 900 psi

K. Remarks/Maintenance record: CALIBRATED TANK SENSOR, REPLACED BREATH HOSE O-RING and SIMULATOR RETURN O-RING, ANNUAL INSPECTION

14 MAY 14
Date
19 May 14 ✓
Date

Charles E. Ed
Field Inspector's Signature
Deb Kashner
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = WFPD 8164.13.00 06/09
05/14/2014 16:37

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 16512080A2
Standard Cyl #? 18
Standard Expiration? 08/01/2014
Oper No? 666666

Flow Cal. Date: 07/01/2011
Slope 685
Intercept -634740

IR Calibration Date: 07/01/2011
 3um 9um

	3um	9um
0th Coef(*100):	-22591	-17234
1st Coef(*100):	265915	132691
2nd Coef(*100):	2841	1562
H2O adj(mg/l*10k):	559	425

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = WFPD 8164.13.00 06/09
05/14/2014 16:54

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:55
02 Std. Sol.	0.018	16:56
03 Room Air	0.000	16:56
04 Std. Sol.	0.019	16:57
05 Room Air	0.000	16:58
06 Std. Sol.	0.020	16:58
07 Room Air	0.000	16:59


08 Sim Temp = 34.0°C

Simul Ser No = DR5115

Std Sol No = 12111

County = 09

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = WFPD 8164.13.00 06/09
05/14/2014 17:00

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:00
02 Std. Sol.	0.300	17:01
03 Room Air	0.000	17:01
04 Std. Sol.	0.304	17:02
05 Room Air	0.000	17:03
06 Std. Sol.	0.305	17:03
07 Room Air	0.000	17:04


08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 13080

County = 09

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC

0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = WFPD 8164.13.00 06/09
05/14/2014 17:05

Test	AC	Time
01 Room Air	0.000	17:06
02 *Subject Test	INT*	17:07
03 Room Air	0.000	17:07

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 16512080A2
Cyl No = 18
Expiration Date = 08/01/2014
County = 09 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

INTERFERENT CHECK
0.100AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = WFPD 8164.13.00 06/09
05/14/2014 17:07

Test	AC	Time
01 Diagnostic	OK	17:08
02 Room Air	RFI*	17:08
03 Room Air	0.000	17:09

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 16512080A2
Cyl No = 18
Expiration Date = 08/01/2014
County = 09 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = WFPD 8164.13.00 06/09
05/14/2014 17:09

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:10
02 Std. Gas	0.081	17:10
03 Room Air	0.000	17:11
04 Std. Gas	0.081	17:11
05 Room Air	0.000	17:12
06 Std. Gas	0.081	17:12
07 Room Air	0.000	17:12

Lot No = 16512080A2
Cyl No = 18
Exp Date = 08/01/2014
County = 09 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = WFPD 8164.13.00 06/09
05/14/2014 17:13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:13
02 Std. Gas	0.081	17:13
03 Room Air	0.000	17:14
04 Std. Gas	0.082	17:14
05 Room Air	0.000	17:15
06 Std. Gas	0.081	17:15
07 Room Air	0.000	17:16

Lot No = 16512080A2
Cyl No = 18
Exp Date = 08/01/2014
County = 09 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
 0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = WFPD 8164.13.00 06/09
05/14/2014 17:16

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:16
02 Std. Gas	0.081	17:17
03 Room Air	0.000	17:17
04 Std. Gas	0.083	17:18
05 Room Air	0.000	17:18
06 Std. Gas	0.081	17:18
07 Room Air	0.000	17:19

Lot No = 16512080A2
Cyl No = 18
Exp Date = 08/01/2014
County = 09 Oper No. = 666666


Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
0.080 AC