

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-005937

Location: TOXL

- | | Check When Done |
|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-----------------|
| A. Is the breath tube heated? | <u>✓</u> |
| B. Is the diagnostic check complete? | <u>✓</u> |
| C. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| D. Print test. (Attach test record.) | <u>✓</u> |
| E. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3377</u> Lot No. <u>13270</u> Rep AC <u>0.020</u> | <u>✓</u> |
| F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5190</u> Lot No. <u>12105</u> Rep AC <u>0.400</u> | <u>✓</u> |
| G. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR5135</u> Lot No. <u>ICS#2</u> Rep AC <u>0.100AC + 0.05% ACETONE</u> | <u>✓</u> |
| H. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. <u>1651208042</u>	Cyl No. <u>019</u>	Exp. Date <u>08/01/2014</u>
Test 1 <u>0.081</u>	Test 4 <u>0.082</u>	Test 7 <u>0.082</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.082</u>
Test 3 <u>0.082</u>	Test 6 <u>0.082</u>	Test 9 <u>0.082</u>
Average <u>0.082</u>		

J. Tank pressure, Level 3 Func D Sub G P: 799 psi Reg 800 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

19 MAY 14
Date

19 May 14
Date

Charles E. Eda
Field Inspector's Signature

Deb Kashner
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.13.00 06/09
05/19/2014 14:09

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 16512080A2
Standard Cyl #? 19
Standard Expiration? 08/01/2014
Oper No? 666666

Flow Cal. Date: 07/19/2013
Slope 683
Intercept -742419

IR Calibration Date: 07/19/2013

	3um	9um
0th Coef(*100):	-21480	-18589
1st Coef(*100):	274353	136859
2nd Coef(*100):	2162	1226
H2O adj(mg/l*10k):	552	490

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.13.00 06/09
05/19/2014 14:20

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:21
02 Std. Sol.	0.017	14:21
03 Room Air	0.000	14:22
04 Std. Sol.	0.018	14:23
05 Room Air	0.000	14:23
06 Std. Sol.	0.018	14:24
07 Room Air	0.000	14:24

08 Sim Temp = 34.0°C

Simul Ser No = DR3377
Std Sol No = 13270
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC*
 0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.13.00 06/09
05/19/2014 14:25

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:26
02 Std. Sol.	0.404	14:27
03 Room Air	0.000	14:27
04 Std. Sol.	0.402	14:28
05 Room Air	0.000	14:28
06 Std. Sol.	0.402	14:29
07 Room Air	0.000	14:30


08 Sim Temp = 34.0°C

Simul Ser No = DR5190

Std Sol No = 12105

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC

0.400 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.13.00 06/09
05/19/2014 14:37

Test	AC	Time
01 Diagnostic	OK	14:38
02 Room Air	RFI*	14:38
03 Room Air	0.000	14:39

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 16512080A2
Cyl No = 19
Expiration Date = 08/01/2014
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.13.00 06/09
05/19/2014 14:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:40
02 Std. Gas	0.081	14:41
03 Room Air	0.000	14:41
04 Std. Gas	0.081	14:42
05 Room Air	0.000	14:42
06 Std. Gas	0.082	14:42
07 Room Air	0.000	14:43

Lot No = 16512080A2
Cyl No = 19
Exp Date = 08/01/2014
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.13.00 06/09
05/19/2014 14:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:44
02 Std. Gas	0.082	14:44
03 Room Air	0.000	14:45
04 Std. Gas	0.081	14:45
05 Room Air	0.000	14:45
06 Std. Gas	0.082	14:46
07 Room Air	0.000	14:46

Lot No = 16512080A2
Cyl No = 19
Exp Date = 08/01/2014
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
D.OBOAC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005937
Location = TOXL 8164.13.00 06/09
05/19/2014 14:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:47
02 Std. Gas	0.082	14:48
03 Room Air	0.000	14:48
04 Std. Gas	0.082	14:49
05 Room Air	0.000	14:49
06 Std. Gas	0.082	14:50
07 Room Air	0.000	14:50

Lot No = 16512080A2
Cyl No = 19
Exp Date = 08/01/2014
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000