

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-005950

Location: FTOT (FT. TOTTEN)

- |   | Check When Done |
|---|-----------------|
| A. Is the breath tube heated?   | <u>✓</u>        |
| B. Is the diagnostic check complete?  | <u>✓</u>        |
| C. Is the time, date, and year correct (re-set if necessary)?   | <u>✓</u>        |
| D. Print test. (Attach test record.)  | <u>✓</u>        |
| E. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5115</u> Lot No <u>12111</u> Rep AC <u>0.020</u>  | <u>✓</u>        |
| F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5143</u> Lot No <u>13080</u> Rep AC <u>0.300</u>  | <u>✓</u>        |
| G. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR3375</u> Lot No <u>ICS#2</u> Rep AC <u>0.100 AC + 0.05% ACETONE</u> | <u>✓</u>        |
| H. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)   | <u>✓</u>        |
| I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)   |                 |

Lot No. 33913080A4 Cyl No. 046 Exp. Date 01/01/2016  
Test 1 0.081 Test 4 0.082 Test 7 0.080  
Test 2 0.081 Test 5 0.081 Test 8 0.081  
Test 3 0.082 Test 6 0.081 Test 9 0.081  
Average 0.081

J. Tank pressure, Level 3 Func D Sub G P: 1101 psi Reg 1000+ psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

15 MAY 14  
Date

19 May 14 ✓  
Date

Charles E. Edin  
Field Inspector's Signature

Deb Kashner  
Reviewed by

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005950  
Location = FTOT                    8164.13.00 06/09  
05/15/2014                            12:38

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#%&^\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?           1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            33913080A4  
Standard Cyl #?            46  
Standard Expiration?        01/01/2016  
Oper No?                    666666

Flow Cal. Date:            06/20/2013  
Slope                      713  
Intercept                  -723515

IR Calibration Date:        06/20/2013  
                                        3um                    9um

-----  
0th Coef(\*100):            -37081                    -23816  
1st Coef(\*100):            290549                    133411  
2nd Coef(\*100):            552                        1090  
H2O adj(mg/l\*10k):        683                        502

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
CHARLES EDER

Remarks:

PRINTTEST

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005950  
Location = FTOT      8164.13.00 06/09  
05/15/2014      12:52

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:53
02 Std. Sol.	0.018	12:54
03 Room Air	0.000	12:54
04 Std. Sol.	0.019	12:55
05 Room Air	0.000	12:56
06 Std. Sol.	0.019	12:56
07 Room Air	0.000	12:57


08 Sim Temp = 34.0°C

Simul Ser No = DR5115

Std Sol No = 12111

County = 62

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks:

Low AC  
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005950  
Location = FTOT      8164.13.00 06/09  
05/15/2014      12:57

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:58
02 Std. Sol.	0.298	12:59
03 Room Air	0.000	12:59
04 Std. Sol.	0.301	13:00
05 Room Air	0.000	13:01
06 Std. Sol.	0.303	13:01
07 Room Air	0.000	13:02

08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 13080

County = 62

Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *HIGH AC*  
*0.300 AC*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005950  
Location = FTOT      8164.13.00 06/09  
05/15/2014      12:42

Test	AC	Time
01 Room Air	0.000	12:43
02 *Subject Test	INT*	12:43
03 Room Air	0.000	12:44

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = INTERFERENT  
Dr. Lic. = ND/DIS821456  
Lot No = 33913080A4  
Cyl No = 46  
Expiration Date = 01/01/2016  
County = 62      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: *INTERFERENT CHECK*  
*0.100 AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-005950  
Location = FTOT                    8164.13.00 06/09  
05/15/2014                            12:40

Test	AC	Time
01 Diagnostic	OK	12:41
02 Room Air	RFI*	12:41
03 Room Air	0.000	12:41

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CHECK  
Dr. Lic. = ND/DIS821456  
Lot No = 33913080A4  
Cyl No = 46  
Expiration Date = 01/01/2016  
County = 62                            Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
CHARLES EDER

Remarks:

*RFI CHECK*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005950  
Location = FTOT      8164.13.00 06/09  
05/15/2014      12:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:46
02 Std. Gas	0.081	12:46
03 Room Air	0.000	12:47
04 Std. Gas	0.081	12:47
05 Room Air	0.000	12:47
06 Std. Gas	0.082	12:48
07 Room Air	0.000	12:48

Lot No = 33913080A4  
Cyl No = 46  
Exp Date = 01/01/2016  
County = 62      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION CHECK  
0.080AC

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-005950  
Location = FTOT                    8164.13.00 06/09  
05/15/2014                            12:49

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:49
02 Std. Gas	0.082	12:50
03 Room Air	0.000	12:50
04 Std. Gas	0.081	12:51
05 Room Air	0.000	12:51
06 Std. Gas	0.081	12:51
07 Room Air	0.000	12:52

Lot No = 33913080A4  
Cyl No = 46  
Exp Date = 01/01/2016  
County = 62                            Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION  
CHECK  
0.080 AC  
Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-005950  
Location = FTOT      8164.13.00 06/09  
05/15/2014      13:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:03
02 Std. Gas	0.080	13:04
03 Room Air	0.000	13:04
04 Std. Gas	0.081	13:04
05 Room Air	0.000	13:05
06 Std. Gas	0.081	13:05
07 Room Air	0.000	13:06

Lot No = 33913080A4  
Cyl No = 46  
Exp Date = 01/01/2016  
County = 62      Oper No. = 666666



Operator Signature  
CHARLES EDER

Remarks: CALIBRATION  
CHECK  
0.080AC

Form 106-I8000