



ETHANOL GAS STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <u>Deb Keshur</u>		
Location <u>Renville Co SO (Mohl)</u>	Intoxilyzer® Serial Number <u>80-004939</u>	
Gas Lot Number ^{DK} <u>29614080A2</u>	Gas Cylinder Number <u>1</u>	Gas Expiration Date <u>12/05/2016</u>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. 0.080 AC
 B. 0.080 AC
 C. 0.019 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- N/A 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <u>Deb Keshur</u>	Date <u>19 May 15</u> ✓
Reviewed By (Crime Laboratory Use Only) <u>[Signature]</u>	Date <u>20 MAY 15</u>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = MOHL 8164.13.00 06/09
05/19/2015 12:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:03
02 Std. Gas	0.080	12:04
03 Room Air	0.000	12:04
04 Std. Gas	0.080	12:05
05 Room Air	0.000	12:05
06 Std. Gas	0.079	12:05
07 Room Air	0.000	12:06

Lot No = 29614080A2
Cyl No = 1
Exp Date = 12/05/2016
County = 38 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC