

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004944  
~~860015~~

Location: West Fargo PD

Check When Done

- A. Is the breath tube heated? ✓
- B. Is the diagnostic check complete? ✓
- C. Is the time, date, and year correct (re-set if necessary)? ✓
- D. Print test. (Attach test record.) ✓
- E. Low AC. Use  $\leq 0.03$  AC in ACA mode. (Attach test record.)  
Sim SN DR5191 Lot No. 15160 Rep AC 0.020 ✓
- F. High AC. Use  $\geq 0.25$  AC in ACA mode. (Attach test record.)  
Sim SN DR3847 Lot No. 16060 Rep AC 0.300 ✓
- G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.  
(Attach test record.)  
Sim SN DR5190 Lot No. ICS 4 Solution 0.10 Acet 0.05% ✓  
*acetone*
- H. RFI check. Run CMS mode. Key radio during CMS test.  
(Attach test record.) ✓
- I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.  
(Attach test record.) 0.080 AC

Lot No. <u>14214080A6</u>	Cyl No. <u>59</u>	Exp. Date <u>07/05/2016</u>
Test 1 <u>0.082</u>	Test 4 <u>0.083</u>	Test 7 <u>0.081</u>
Test 2 <u>0.082</u>	Test 5 <u>0.083</u>	Test 8 <u>0.083</u>
Test 3 <u>0.082</u>	Test 6 <u>0.082</u>	Test 9 <u>0.083</u>
Average <u>0.082</u>		

J. Tank pressure, Level 3 Func D Sub G P: 459 psi Reg ~450 psi

K. Remarks/Maintenance record: Annual inspection, replaced breath hose & simulator return o-ring.

18 May 16  
Date  
19 MAY 2016 ✓  
Date

Deb Shanaver  
Field Inspector's Signature  
[Signature]  
Reviewed by



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004944  
Location = WFPD      8164.13.00 06/09  
05/18/2016      10:37

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:38
02 Std. Sol.	0.019	10:39
03 Room Air	0.000	10:39
04 Std. Sol.	0.018	10:40
05 Room Air	0.000	10:40
06 Std. Sol.	0.020	10:41
07 Room Air	0.000	10:42

08 Sim Temp = 34.0°C

Simul Ser No = DR5191  
Std Sol No = 15160  
County = ~~09-DS~~  
          39      Oper No. = 777777

*Deb Shawaver*

Operator Signature  
DEB KASHUR

Remarks:

*Low alcohol 0.020 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004944  
Location = WFPD      8164.13.00 06/09  
05/18/2016      10:31

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:32
02 Std. Sol.	0.311	10:33
03 Room Air	0.000	10:33
04 Std. Sol.	0.310	10:34
05 Room Air	0.000	10:35
06 Std. Sol.	0.311	10:35
07 Room Air	0.000	10:36

08 Sim Temp = 34.0°C

Simul Ser No = DR3847

Std Sol No = 16060

County = ~~09~~ <sup>DS</sup>  
39

Oper No. = 777777

*Deb Shanaver*

Operator Signature  
DEB KASHUR

Remarks:

*High alcohol      0.300 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004944  
Location = WFPD                    8164.13.00 06/09  
05/18/2016                            10:29

Test	AC	Time
01 Room Air	0.000	10:30
02 *Subject Test	INT*	10:30
03 Room Air	0.000	10:31

\*Invalid Test  
Interferent Detected

Sub Name = INTERFERENT, NA NA  
Sub DOB = 08/23/1987  
Sub Sex = Unknown                    Weight = NA  
Test = DUI                              Cit = NA  
Dr. Lic. = ND/NA  
Lot No = 14214080A6  
Cyl No = 59  
Expiration Date = 07/05/2016  
County = ~~09-DS~~                    Oper No. = 777777  
          39

Deb Shanaver

Operator Signature  
DEB KASHUR

Remarks:

*Interferent detected 0.10AC + 0.05% acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004944  
Location = WFPD                    8164.13.00 06/09  
05/18/2016                            10:27

Test	AC	Time
01 Diagnostic	OK	10:27
02 Room Air	RFI*	10:28
03 Room Air	0.000	10:28

\*Invalid Test  
Inhibited - RFI

Sub Name = RFI, NA NA  
Sub DOB = 08/23/1978  
Sub Sex = Unknown                    Weight = NA  
Test = DUI                              Cit = NA  
Dr. Lic. = ND/NA  
Lot No = 14214080A6  
Cyl No = 59  
Expiration Date = 07/05/2016  
County = ~~09-D5~~                    Oper No. = 777777  
          39

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Shawver

Operator Signature  
DEB KASHUR

Remarks:

RFI test


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004944  
Location = WFPD                    8164.13.00 06/09  
05/18/2016                            10:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:15
02 Std. Gas	0.082	10:16
03 Room Air	0.000	10:16
04 Std. Gas	0.082	10:16
05 Room Air	0.000	10:17
06 Std. Gas	0.082	10:17
07 Room Air	0.000	10:18

Lot No = 14214080A6  
Cyl No = 59  
Exp Date = 07/05/2016  
County = ~~09 DS~~ 39                    Oper No. = 777777

  
\_\_\_\_\_  
Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004944  
Location = WFPD      8164.13.00 06/09  
05/18/2016      10:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:19
02 Std. Gas	0.083	10:19
03 Room Air	0.000	10:19
04 Std. Gas	0.083	10:20
05 Room Air	0.000	10:20
06 Std. Gas	0.082	10:21
07 Room Air	0.000	10:21

Lot No = 14214080A6  
Cyl No = 59  
Exp Date = 07/05/2016  
County = ~~09 DS~~  
          39      Oper No. = 777777

Deb Shanaver  
Operator Signature  
DEB KASHUR

Remarks:

Calibration check 0.080 AC




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004944  
Location = WFPD      8164.13.00 06/09  
05/18/2016      10:23

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:24
02 Std. Gas	0.081	10:24
03 Room Air	0.000	10:25
04 Std. Gas	0.083	10:25
05 Room Air	0.000	10:25
06 Std. Gas	0.083	10:26
07 Room Air	0.000	10:26

Lot No = 14214080A6  
Cyl No = 59  
Exp Date = 07/05/2016  
County = ~~09 DS~~  
          39      Oper No. = 777777

  
\_\_\_\_\_  
Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*  
Form 106-I8000