

INTOXILYZER® 8000 INSPECTION

Serial No.: 8D-004947

Location: Logan Co SO

- | | Check When Done |
|--|-----------------|
| A. Is the breath tube heated? | <u>✓</u> |
| B. Is the diagnostic check complete? | <u>✓</u> |
| C. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| D. Print test. (Attach test record.) | <u>✓</u> |
| E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.) Sim SN <u>DR5191</u> Lot No. <u>15160</u> Rep AC <u>0.020</u> | <u>✓</u> |
| F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.) Sim SN <u>DR 3847</u> Lot No. <u>16060</u> Rep AC <u>0.300</u> | <u>✓</u> |
| G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode. (Attach test record.) Sim SN <u>DR5190</u> Lot No. <u>ICS4</u> Solution <u>0.10AC + 0.05% acetone</u> | <u>✓</u> |
| H. RFI check. Run CMS mode. Key radio during CMS test. (Attach test record.) | <u>✓</u> |
| I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode. (Attach test record.) <u>0.080</u> AC | |

| | | |
|---------------------------|---------------------|-----------------------------|
| Lot No. <u>29614080A2</u> | Cyl No. <u>31</u> | Exp. Date <u>12/05/2016</u> |
| Test 1 <u>0.081</u> | Test 4 <u>0.082</u> | Test 7 <u>0.080</u> |
| Test 2 <u>0.081</u> | Test 5 <u>0.081</u> | Test 8 <u>0.082</u> |
| Test 3 <u>0.080</u> | Test 6 <u>0.080</u> | Test 9 <u>0.081</u> |
| Average <u>0.081</u> | | |

J. Tank pressure, Level 3 Func D Sub G P: 995 psi Reg 1000 psi

K. Remarks/Maintenance record: Annual inspection, replaced breath hose & simulator return O-rings

17 May 16
Date

19 MAY 2016 ✓
Date

Deb Shanaver
Field Inspector's Signature

[Signature]
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = NAPO 8164.13.00 06/09
05/17/2016 10:36

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 29614080A2
Standard Cyl #? 31
Standard Expiration? 12/05/2016
Oper No? 777777

Flow Cal. Date: 01/19/2010
Slope 679
Intercept -681178

IR Calibration Date: 05/12/2015

| | 3um | 9um |
|--------------------|--------|--------|
| 0th Coef(*100): | -14993 | -20474 |
| 1st Coef(*100): | 257346 | 132154 |
| 2nd Coef(*100): | 2870 | 1330 |
| H2O adj(mg/l*10k): | 492 | 531 |

***** Printer Test End *****

Deb Shanaver

Operator Signature
DEB KASHUR

Remarks:

Print test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = NAPO 8164.13.00 06/09
05/17/2016 10:45

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 10:46 |
| 02 Std. Sol. | 0.017 | 10:46 |
| 03 Room Air | 0.000 | 10:47 |
| 04 Std. Sol. | 0.017 | 10:48 |
| 05 Room Air | 0.000 | 10:48 |
| 06 Std. Sol. | 0.017 | 10:49 |
| 07 Room Air | 0.000 | 10:49 |

08 Sim Temp = 34.0°C

Simul Ser No = DR5191
Std Sol No = 15160
County = 24 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

ADS Low alcohol 0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = NAPO 8164.13.00 06/09
05/17/2016 10:54

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 10:55 |
| 02 Std. Sol. | 0.291 | 10:56 |
| 03 Room Air | 0.000 | 10:56 |
| 04 Std. Sol. | 0.293 | 10:57 |
| 05 Room Air | 0.000 | 10:57 |
| 06 Std. Sol. | 0.294 | 10:58 |
| 07 Room Air | 0.000 | 10:59 |

08 Sim Temp = 34.0°C

Simul Ser No = DR3847
Std Sol No = 16060
County = 24 Oper No. = 777777


Operator Signature
DEB KASHUR

Remarks:

High AC 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = NAPO 8164.13.00 06/09
05/17/2016 10:38

| Test | AC | Time |
|------------------|-------|-------|
| 01 Room Air | 0.000 | 10:39 |
| 02 *Subject Test | INT* | 10:40 |
| 03 Room Air | 0.000 | 10:40 |

*Invalid Test
Interferent Detected

Deb Shanaver
Operator Signature
DEB KASHUR

Remarks:

Interferent detected 0.10AC + 0.05%
acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = NAPO 8164.13.00 06/09
05/17/2016 10:42

| Test | AC | Time |
|---------------|-------|-------|
| 01 Diagnostic | OK | 10:43 |
| 02 Room Air | RFI* | 10:43 |
| 03 Room Air | 0.000 | 10:43 |

*Invalid Test
Inhibited - RFI

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Shanover
Operator Signature
DEB KASHUR

Remarks:

RFI ~~test~~^{DS}
test
Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = NAPO 8164.13.00 06/09
05/17/2016 10:21

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 10:22 |
| 02 Std. Gas | 0.081 | 10:22 |
| 03 Room Air | 0.000 | 10:23 |
| 04 Std. Gas | 0.081 | 10:23 |
| 05 Room Air | 0.000 | 10:24 |
| 06 Std. Gas | 0.080 | 10:24 |
| 07 Room Air | 0.000 | 10:24 |

Lot No = 29614080A2
Cyl No = 31
Exp Date = 12/05/2016
County = 24 Oper No. = 777777

Deb Shanaver

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = NAPO 8164.13.00 06/09
05/17/2016 10:25

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 10:25 |
| 02 Std. Gas | 0.082 | 10:25 |
| 03 Room Air | 0.000 | 10:26 |
| 04 Std. Gas | 0.081 | 10:26 |
| 05 Room Air | 0.000 | 10:27 |
| 06 Std. Gas | 0.080 | 10:27 |
| 07 Room Air | 0.000 | 10:28 |

Lot No = 29614080A2
Cyl No = 31
Exp Date = 12/05/2016
County = 24 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004947
Location = NAPO 8164.13.00 06/09
05/17/2016 10:28

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 10:29 |
| 02 Std. Gas | 0.080 | 10:29 |
| 03 Room Air | 0.000 | 10:29 |
| 04 Std. Gas | 0.082 | 10:30 |
| 05 Room Air | 0.000 | 10:30 |
| 06 Std. Gas | 0.081 | 10:31 |
| 07 Room Air | 0.000 | 10:31 |

Lot No = 29614080A2
Cyl No = 31
Exp Date = 12/05/2016
County = 24 Oper No. = 777777

Deb Shanaver
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080AC