



**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59281 (1-2010)

Serial Number: <u>80-004956</u>	Instrument Location: <u>Emmons County Sheriff's Office</u>
Reason for Install/Repair:	
<input type="checkbox"/> Install After Lab Repair/Inspection <input checked="" type="checkbox"/> Other (Specify) <u>NEW</u>	

Check When Done:

- A. Surge Protector Installed/Properly Grounded.
- B. Telephone Line Connected to Intoxilyzer® 8000.
- C. Breath Tube Heated.
- D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- E. Review/Enter Gas Setup (Level 1, Function S).
- F. Print Test (Level 1, Function P). (Attach Test Record.)
- G. ACA Test (Level 1, Function C). (Attach Test Record.)
- H. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- I. Repair and/or Maintenance Performed: Install

B. L. \_\_\_\_\_ 5/20/10 \_\_\_\_\_  
 Field Inspector Signature Date

Deb Kashur \_\_\_\_\_ 01 June '10 \_\_\_\_\_  
 Reviewed By Date

80-004956  
P. 34/39 FM




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004956  
Location = LINT      8164.13.00 06/09  
05/20/2010      13:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:48
02 Std. Gas	0.078	13:49
03 Room Air	0.000	13:49
04 Std. Gas	0.077	13:49
05 Room Air	0.000	13:50
06 Std. Gas	0.077	13:50
07 Room Air	0.000	13:51

Lot No = 709352  
Cyl No = 16  
Exp Date = 12/23/2011  
County = 15      Oper No. = 100630



Operator Signature  
BRIAN M. LURA

Remarks:

*Install*

Form 106-I8000

80-004956  
P. 34/39 *KH*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004956  
Location = LINT      8164.13.00 06/09  
05/20/2010      13:52

Test	AC	Time
01 Diagnostic	OK	13:54
02 Room Air	RFI*	13:54
03 Room Air	0.000	13:54

\*Invalid Test  
Inhibited - RFI

Sub Name = LURA, BRIAN MICHAEL  
Sub DOB = 10/31/1977  
Sub Sex = Male      Weight = 220  
Test = DUI      Cit = NA  
Dr. Lic. = ND/LUR779393  
Lot No = 709352  
Cyl No = 16  
Expiration Date = 12/23/2011  
County = 15      Oper No. = 100630

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
BRIAN M. LURA

Remarks:

Install RFI Test

Form 106-I8000

80-004956  
P. 39/39 KLM