



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Brian Lura</u>		
Location: <u>Kennett County Sheriff's Office</u>		
Intoxilyzer® Serial Number: <u>80-004956</u>		
Gas Lot Number: <u>709352</u>	Gas Cylinder Number: <u>16</u>	Expiration Date: <u>12/23/11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>078</u> %		
Test 2: 0. <u>078</u> %		
Test 3: 0. <u>078</u> %		
(ATTACH TEST RECORD)		

B-L  
Chemical Test Operator Signature

5/20/10  
Date

Deb Kashner  
Reviewed By

01 June '10  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004956  
Location = LINT      8164.13.00 06/09  
05/20/2010      13:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:41
02 Std. Gas	0.078	13:41
03 Room Air	0.000	13:41
04 Std. Gas	0.078	13:42
05 Room Air	0.000	13:42
06 Std. Gas	0.078	13:43
07 Room Air	0.000	13:43

Lot No = 709352  
Cyl No = 16  
Exp Date = 12/23/2011  
County = 15      Oper No. = 100630



Operator Signature  
BRIAN M. LURA

Remarks:

*Install of machine*

Form 106-I8000

80-004956  
P. 35/35 JH