

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004939

Location: Reynolds Co SO (MOHL)

- Check When Done
- A. Is the breath tube heated? ✓
  - B. Is the diagnostic check complete? ✓
  - C. Is the time, date, and year correct (re-set if necessary)? ✓
  - D. Print test. (Attach test record.) ✓
  - E. Low AC. Use  $\leq 0.03$  AC in ACA mode. (Attach test record.)  
Sim SN DR 5143 Lot No. 201501B Rep AC 0.020 ✓
  - F. High AC. Use  $\geq 0.25$  AC in ACA mode. (Attach test record.)  
Sim SN DR 3455 Lot No. 14270 Rep AC 0.300 ✓
  - G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.  
(Attach test record.)  
Sim SN DR 5132 Lot No. OK 1CS 2 ~~162~~ Solution 0.10 AC + 0.05% acetone ✓
  - H. RFI check. Run CMS mode. Key radio during CMS test.  
(Attach test record.) ✓
  - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.  
(Attach test record.) 0.080 AC

Lot No. 29614080 A2 Cyl No. 1 Exp. Date 12/05/2016  
Test 1 0.080 Test 4 0.081 Test 7 0.081  
Test 2 0.079 Test 5 0.080 Test 8 0.080  
Test 3 0.080 Test 6 0.081 Test 9 0.080  
Average 0.080

J. Tank pressure, Level 3 Func D Sub G P: 980 psi Reg 1000 psi

K. Remarks/Maintenance record: Annual inspection, replaced o-rings on breath hose & simulator return, replaced gas cyl.

19 May 15  
Date

20 MAY 15 ✓  
Date

Deb Kasher  
Field Inspector's Signature

Charles E. [Signature]  
Reviewed by

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004939  
Location = MOHL                    8164.13.00 06/09  
05/19/2015                            11:19

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:                    Enabled  
Start Test Sequence:                DABACABA  
Display Prelim Rslt?                Yes  
Display Third Digit?                Yes  
Inhib Printer(Y/N)?                No  
Display Volume?                    No  
Disable On Memfull?                Yes  
# of Print Copies?                  1  
Select Std (D/W/I)?                Dry  
Standard Value?                    0.080  
Standard Lot #?                    29614080A2  
Standard Cyl #?                    1  
Standard Expiration?                12/05/2016  
Oper No?                            777777

Flow Cal. Date:                    01/18/2010  
Slope                                724  
Intercept                            -779308

IR Calibration Date:                03/10/2010  
    3um                                9um

-----  
0th Coef(\*100):                    -45908                                -35439  
1st Coef(\*100):                    263128                                135479  
2nd Coef(\*100):                    1658                                    1038  
H2O adj(mg/l\*10k):                832                                    527

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
DEB KASHUR

Remarks:



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004939  
Location = MOHL      8164.13.00 06/09  
05/19/2015      11:43

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:44
02 Std. Sol.	0.016	11:44
03 Room Air	0.000	11:45
04 Std. Sol.	0.016	11:46
05 Room Air	0.000	11:46
06 Std. Sol.	0.016	11:47
07 Room Air	0.000	11:48

08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 201501B

County = 38

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Low AC 0.020 AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004939  
Location = MOHL      8164.13.00 06/09  
05/19/2015      11:57

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:57
02 Std. Sol.	0.289	11:58
03 Room Air	0.000	11:59
04 Std. Sol.	0.292	11:59
05 Room Air	0.000	12:00
06 Std. Sol.	0.293	12:01
07 Room Air	0.000	12:01

08 Sim Temp = 34.0°C

Simul Ser No = DR3455

Std Sol No = 14270

County = 38

Oper No. = 777777

Deb Kasher

Operator Signature

DEB KASHUR

Remarks:

High AC 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004939  
Location = MOHL      8164.13.00 06/09  
05/19/2015      11:50

Test	AC	Time
01 Room Air	0.000	11:50
02 *Subject Test	INT*	11:51
03 Room Air	0.000	11:51

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Simul Ser No = DR3455  
Std Sol No = 14270  
County = 38      Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Interferent check 0.10 AC + 0.05% acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004939  
Location = MOHL                    8164.13.00 06/09  
05/19/2015                            11:36

Test	AC	Time
01 Diagnostic	OK	11:37
02 Room Air	0.000	11:38
03 *Subject Test	RFI*	11:41
04 Room Air	0.000	11:41

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 29614080A2  
Cyl No = 1  
Expiration Date = 12/05/2016  
County = 38                            Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*RFI check*

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004939  
Location = MOHL      8164.13.00 06/09  
05/19/2015      11:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:21
02 Std. Gas	0.080	11:22
03 Room Air	0.000	11:22
04 Std. Gas	0.079	11:22
05 Room Air	0.000	11:23
06 Std. Gas	0.080	11:23
07 Room Air	0.000	11:24

Lot No = 29614080A2  
Cyl No = 1  
Exp Date = 12/05/2016  
County = 38      Oper No. = 777777

  
\_\_\_\_\_  
Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004939  
Location = MOHL      8164.13.00 06/09  
05/19/2015      11:25

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:25
02 Std. Gas	0.081	11:26
03 Room Air	0.000	11:26
04 Std. Gas	0.080	11:27
05 Room Air	0.000	11:27
06 Std. Gas	0.081	11:27
07 Room Air	0.000	11:28

Lot No = 29614080A2  
Cyl No = 1  
Exp Date = 12/05/2016  
County = 38      Oper No. = 777777

  
\_\_\_\_\_  
Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004939  
Location = MOHL      8164.13.00 06/09  
05/19/2015      11:28

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:29
02 Std. Gas	0.081	11:29
03 Room Air	0.000	11:30
04 Std. Gas	0.080	11:30
05 Room Air	0.000	11:31
06 Std. Gas	0.080	11:31
07 Room Air	0.000	11:32

Lot No = 29614080A2

Cyl No = 1

Exp Date = 12/05/2016

County = 38

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000