

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004954

Location: Rolette Co SO

- Check When Done
- A. Is the breath tube heated? ✓
- B. Is the diagnostic check complete? ✓
- C. Is the time, date, and year correct (re-set if necessary)? ✓
- D. Print test. (Attach test record.) ✓
- E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5143 Lot No. 201501 B Rep AC 0.020 ✓
- F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3455 Lot No. 14270 Rep AC 0.300 ✓
- G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR5132 Lot No. 1CS2 Solution 0.10ACT 0.05% acetone ✓
- H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
- I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 33913080A4 Cyl No. 39 Exp. Date 01/01/2016

Test 1 <u>0.081</u>	Test 4 <u>0.082</u>	Test 7 <u>0.082</u>
Test 2 <u>0.082</u>	Test 5 <u>0.082</u>	Test 8 <u>0.081</u>
Test 3 <u>0.082</u>	Test 6 <u>0.082</u>	Test 9 <u>0.082</u>

Average 0.082

J. Tank pressure, Level 3 Func D Sub G P: 917 psi Reg 900 psi

K. Remarks/Maintenance record: Annual inspection

19 May 15
Date

20 MAY 15 ✓
Date

D. b. Kashner
Field Inspector's Signature

[Signature]
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = ROLL 8164.13.00 06/09
05/19/2015 15:40

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 33913080A4
Standard Cyl #? 39
Standard Expiration? 01/01/2016
Oper No? 777777

Flow Cal. Date: 08/27/2012
Slope 670
Intercept -570610

IR Calibration Date: 08/27/2012
 3um 9um

0th Coef(*100): -22689 -21316
1st Coef(*100): 260406 132331
2nd Coef(*100): 1084 919
H2O adj(mg/l*10k): 570 470

***** Printer Test End *****



Operator Signature
DEB KASHUR

Remarks:



Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = ROLL 8164.13.00 06/09
05/19/2015 15:35

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:36
02 Std. Sol.	0.019	15:37
03 Room Air	0.000	15:37
04 Std. Sol.	0.020	15:38
05 Room Air	0.000	15:38
06 Std. Sol.	0.021	15:39
07 Room Air	0.000	15:40

08 Sim Temp = 34.0°C

Simul Ser No = DR5143
Std Sol No = 201501B
County = 40 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Low AC 0.020Ac

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = ROLL 8164.13.00 06/09
05/19/2015 15:30

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:31
02 Std. Sol.	0.295	15:31
03 Room Air	0.000	15:32
04 Std. Sol.	0.294	15:33
05 Room Air	0.000	15:33
06 Std. Sol.	0.294	15:34
07 Room Air	0.000	15:34

08 Sim Temp = 34.0°C

Simul Ser No = DR3455
Std Sol No = 14270
County = 40 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

High AC 0.300 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = ROLL 8164.13.00 06/09
05/19/2015 15:28

Test	AC	Time
01 Room Air	0.000	15:28
02 *Subject Test	INT*	15:29
03 Room Air	0.000	15:29

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 33913080A4
Cyl No = 39
Expiration Date = 01/01/2016
County = 40 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.104e + 0.05%
acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = ROLL 8164.13.00 06/09
05/19/2015 15:26

Test	AC	Time
01 Diagnostic	OK	15:27
02 Room Air	RFI*	15:27
03 Room Air	0.000	15:27

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 33913080A4
Cyl No = 39
Expiration Date = 01/01/2016
County = 40 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB KASHUR

Remarks:

RFI check

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = ROLL 8164.13.00 06/09
05/19/2015 15:16

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:16
02 Std. Gas	0.081	15:17
03 Room Air	0.000	15:17
04 Std. Gas	0.082	15:18
05 Room Air	0.000	15:18
06 Std. Gas	0.082	15:18
07 Room Air	0.000	15:19

Lot No = 33913080A4
Cyl No = 39
Exp Date = 01/01/2016
County = 40 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = ROLL 8164.13.00 06/09
05/19/2015 15:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:20
02 Std. Gas	0.082	15:20
03 Room Air	0.000	15:21
04 Std. Gas	0.082	15:21
05 Room Air	0.000	15:21
06 Std. Gas	0.082	15:22
07 Room Air	0.000	15:22

Lot No = 33913080A4
Cyl No = 39
Exp Date = 01/01/2016
County = 40 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080AE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = ROLL 8164.13.00 06/09
05/19/2015 15:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:23
02 Std. Gas	0.082	15:23
03 Room Air	0.000	15:24
04 Std. Gas	0.081	15:24
05 Room Air	0.000	15:25
06 Std. Gas	0.082	15:25
07 Room Air	0.000	15:26

Lot No = 33913080A4
Cyl No = 39
Exp Date = 01/01/2016
County = 40 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080AC

Form 106-I8000