

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004948

Location: STAN (MOUNT TRAIL Co. S.O.)

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5143 Lot No. 201501B Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3455 Lot No. 14270 Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR5132 Lot No. ICS #2 Solution 0.10 AC + 0.05% ACE ✓
 - H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

| | | |
|---------------------------|---------------------|-------------------------|
| Lot No. <u>13314080A2</u> | Cyl No. <u>54</u> | Exp. Date <u>6/5/16</u> |
| Test 1 <u>0.080</u> | Test 4 <u>0.080</u> | Test 7 <u>0.081</u> |
| Test 2 <u>0.080</u> | Test 5 <u>0.080</u> | Test 8 <u>0.080</u> |
| Test 3 <u>0.081</u> | Test 6 <u>0.081</u> | Test 9 <u>0.080</u> |
| Average <u>0.080</u> | | |

J. Tank pressure, Level 3 Func D Sub G P: 1015 psi Reg 1000+ psi

K. Remarks/Maintenance record: REPLACED SIMULATOR RETURN O-RINGS

ANNUAL INSPECTION

21 MAY 15
Date

Charles E. Edler
Field Inspector's Signature

21 May 15
Date

Deb Kashner
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = STAN 8164.13.00 06/09
05/21/2015 11:15

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 11:15 |
| 02 Std. Sol. | 0.018 | 11:16 |
| 03 Room Air | 0.000 | 11:17 |
| 04 Std. Sol. | 0.019 | 11:17 |
| 05 Room Air | 0.000 | 11:18 |
| 06 Std. Sol. | 0.019 | 11:19 |
| 07 Room Air | 0.000 | 11:19 |


08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 201501B

County = 31

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = STAN 8164.13.00 06/09
05/21/2015 11:25

WET CAL CHECK

| Test | AC | Time |
|--------------|-------|-------|
| 01 Room Air | 0.000 | 11:26 |
| 02 Std. Sol. | 0.301 | 11:27 |
| 03 Room Air | 0.000 | 11:27 |
| 04 Std. Sol. | 0.301 | 11:28 |
| 05 Room Air | 0.000 | 11:28 |
| 06 Std. Sol. | 0.300 | 11:29 |
| 07 Room Air | 0.000 | 11:30 |

08 Sim Temp = 34.0°C

Simul Ser No = DR3455

Std Sol No = 14270

County = 31

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC

0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = STAN 8164.13.00 06/09
05/21/2015 11:11

| Test | AC | Time |
|------------------|-------|-------|
| 01 Room Air | 0.000 | 11:12 |
| 02 *Subject Test | INT* | 11:12 |
| 03 Room Air | 0.000 | 11:13 |

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 13314080A2
Cyl No = 54
Expiration Date = 06/05/2016
County = 31 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check

0.10 AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = STAN 8164.13.00 06/09
05/21/2015 10:59

| Test | AC | Time |
|------------------|-------|-------|
| 01 Diagnostic | OK | 11:00 |
| 02 Room Air | 0.000 | 11:00 |
| 03 *Subject Test | RFI* | 11:01 |
| 04 Room Air | 0.000 | 11:01 |

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 13314080A2
Cyl No = 54
Expiration Date = 06/05/2016
County = 31 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = STAN 8164.13.00 06/09
05/21/2015 11:02

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 11:03 |
| 02 Std. Gas | 0.080 | 11:03 |
| 03 Room Air | 0.000 | 11:04 |
| 04 Std. Gas | 0.080 | 11:04 |
| 05 Room Air | 0.000 | 11:04 |
| 06 Std. Gas | 0.081 | 11:05 |
| 07 Room Air | 0.000 | 11:05 |

Lot No = 13314080A2
Cyl No = 54
Exp Date = 06/05/2016
County = 31 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = STAN 8164.13.00 06/09
05/21/2015 11:07

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 11:07 |
| 02 Std. Gas | 0.080 | 11:07 |
| 03 Room Air | 0.000 | 11:08 |
| 04 Std. Gas | 0.080 | 11:08 |
| 05 Room Air | 0.000 | 11:09 |
| 06 Std. Gas | 0.081 | 11:09 |
| 07 Room Air | 0.000 | 11:10 |

Lot No = 13314080A2
Cyl No = 54
Exp Date = 06/05/2016
County = 31 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004948
Location = STAN 8164.13.00 06/09
05/21/2015 11:32

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 11:32 |
| 02 Std. Gas | 0.081 | 11:33 |
| 03 Room Air | 0.000 | 11:33 |
| 04 Std. Gas | 0.080 | 11:34 |
| 05 Room Air | 0.000 | 11:34 |
| 06 Std. Gas | 0.080 | 11:34 |
| 07 Room Air | 0.000 | 11:35 |

Lot No = 13314080A2
Cyl No = 54
Exp Date = 06/05/2016
County = 31 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000