

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004935

Location: Ward 6 Corr

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5115 Lot No. 12111 Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5143 Lot No. 13080 Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3375 Lot No. 105#2 Rep AC 0.10 AC + 0.05% acetone ✓
 - H. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 1651208042 Cyl No. 15 Exp. Date 08/01/2014
Test 1 0.079 Test 4 0.078 Test 7 0.077
Test 2 0.079 Test 5 0.077 Test 8 0.078
Test 3 0.078 Test 6 0.077 Test 9 0.078
Average 0.078

J. Tank pressure, Level 3 Func D Sub G P: 877 psi Reg 875 psi

K. Remarks/Maintenance record: annual inspection

21 May 14
Date

22 MAY 14 ✓
Date

Deb Kashur
Field Inspector's Signature

Charles E. Edin
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = WARD 8164.13.00 06/09
05/21/2014 09:02

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 16512080A2
Standard Cyl #? 15
Standard Expiration? 08/01/2014
Oper No? 777777

Flow Cal. Date: 07/12/2012
Slope 655
Intercept -510289

IR Calibration Date: 04/23/2013
 3um 9um

0th Coef(*100): -11131 -13952
1st Coef(*100): 263852 129376
2nd Coef(*100): 3246 1683
H2O adj(mg/l*10k): 464 535

***** Printer Test End *****



Operator Signature
DEB KASHUR

Remarks:



Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = WARD 8164.13.00 06/09
05/21/2014 08:52

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:52
02 Std. Sol.	0.017	08:53
03 Room Air	0.000	08:54
04 Std. Sol.	0.019	08:54
05 Room Air	0.000	08:55
06 Std. Sol.	0.019	08:55
07 Room Air	0.000	08:56

08 Sim Temp = 34.0°C

Simul Ser No = DR5115

Std Sol No = 12111

County = 51

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Low AC 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = WARD 8164.13.00 06/09
05/21/2014 08:57

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:57
02 Std. Sol.	0.291	08:59
03 Room Air	0.000	08:59
04 Std. Sol.	0.295	09:00
05 Room Air	0.000	09:00
06 Std. Sol.	0.296	09:01
07 Room Air	0.000	09:02

08 Sim Temp = 34.0°C

Simul Ser No = DR5143
Std Sol No = 13080
County = 51 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

High AC 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = WARD 8164.13.00 06/09
05/21/2014 09:12

Test	AC	Time
01 Room Air	0.000	09:12
02 *Subject Test	INT*	09:13
03 Room Air	0.000	09:13

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 15
Expiration Date = 08/01/2014
County = 51 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.10AC + 0.05% acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = WARD 8164.13.00 06/09
05/21/2014 09:14

Test	AC	Time
01 Diagnostic	OK	09:15
02 Room Air	RFI*	09:15
03 Room Air	0.000	09:15

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 15
Expiration Date = 08/01/2014
County = 51 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB KASHUR

Remarks:

RF 1 check

Form 106-18000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = WARD 8164.13.00 06/09
05/21/2014 08:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:49
02 Std. Gas	0.079	08:49
03 Room Air	0.000	08:50
04 Std. Gas	0.079	08:50
05 Room Air	0.000	08:50
06 Std. Gas	0.078	08:51
07 Room Air	0.000	08:51

Lot No = 16512080A2
Cyl No = 15
Exp Date = 08/01/2014
County = 51 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

to^{DK} Calibration check 0.080AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = WARD 8164.13.00 06/09
05/21/2014 09:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:16
02 Std. Gas	0.078	09:16
03 Room Air	0.000	09:17
04 Std. Gas	0.077	09:17
05 Room Air	0.000	09:18
06 Std. Gas	0.077	09:18
07 Room Air	0.000	09:19

Lot No = 16512080A2
Cyl No = 15
Exp Date = 08/01/2014
County = 51 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = WARD 8164.13.00 06/09
05/21/2014 09:23

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:24
02 Std. Gas	0.077	09:24
03 Room Air	0.000	09:25
04 Std. Gas	0.078	09:25
05 Room Air	0.000	09:26
06 Std. Gas	0.078	09:26
07 Room Air	0.000	09:27

Lot No = 16512080A2
Cyl No = 15
Exp Date = 08/01/2014
County = 51 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080AC