

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004939

Location: MOHL - Renville Co SD

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5115 Lot No. 12111 Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5143 Lot No. 13080 Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3375 Lot No. 1CS#2 Rep AC 0.10AC + 0.05% acetone ✓
 - H. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 16512080A2 Cyl No. 8 Exp. Date 08/01/2014

Test 1 <u>0.078</u>	Test 4 <u>0.080</u>	Test 7 <u>0.080</u>
Test 2 <u>0.079</u>	Test 5 <u>0.079</u>	Test 8 <u>0.079</u>
Test 3 <u>0.079</u>	Test 6 <u>0.078</u>	Test 9 <u>0.078</u>

Average 0.079

J. Tank pressure, Level 3 Func D Sub G P: 879 psi Reg 880 psi

K. Remarks/Maintenance record: Annual inspection

20 May 14
Date

22 MAY 14 ✓
Date

Deb Kashur
Field Inspector's Signature

Charles E. Edin
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = MOHL 8164.13.00 06/09
05/20/2014 14:47

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890==|
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890==|
ABCDEFGHIJKLMN0PQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 16512080A2
Standard Cyl #? 8
Standard Expiration? 08/01/2014
Oper No? 777777

Flow Cal. Date: 01/18/2010
Slope 724
Intercept -779308

IR Calibration Date: 03/10/2010

	3um	9um
0th Coef(*100):	-45908	-35439
1st Coef(*100):	263128	135479
2nd Coef(*100):	1658	1038
H2O adj(mg/l*10k):	832	527

***** Printer Test End *****



Operator Signature
DEB KASHUR

Remarks:



Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = MOHL 8164.13.00 06/09
05/20/2014 15:05

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:06
02 Std. Sol.	0.018	15:07
03 Room Air	0.000	15:07
04 Std. Sol.	0.019	15:08
05 Room Air	0.000	15:09
06 Std. Sol.	0.019	15:09
07 Room Air	0.000	15:10

08 Sim Temp = 34.0°C

Simul Ser No = DR5115
Std Sol No = 12111
County = 38 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Low Ac 0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = MOHL 8164.13.00 06/09
05/20/2014 15:13

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:14
02 Std. Sol.	0.312	15:15
03 Room Air	0.000	15:15
04 Std. Sol.	0.307	15:16
05 Room Air	0.000	15:17
06 Std. Sol.	0.305	15:17
07 Room Air	0.000	15:18

08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 13080

County = 38

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

High AC 0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = MOHL 8164.13.00 06/09
05/20/2014 14:59

Test	AC	Time
01 Room Air	0.000	14:59
02 *Subject Test	INT*	15:00
03 Room Air	0.000	15:01

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 8
Expiration Date = 08/01/2014
County = 38 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

*Interferent check 0.100AC +
0.05% acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = MOHL 8164.13.00 06/09
05/20/2014 15:02

Test	AC	Time
01 Diagnostic	OK	15:03
02 Room Air	RFI*	15:03
03 Room Air	0.000	15:04

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 8
Expiration Date = 08/01/2014
County = 38 Oper No. = 77777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:
RFI check

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = MOHL 8164.13.00 06/09
05/20/2014 14:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:48
02 Std. Gas	0.078	14:49
03 Room Air	0.000	14:49
04 Std. Gas	0.079	14:50
05 Room Air	0.000	14:50
06 Std. Gas	0.079	14:51
07 Room Air	0.000	14:51

Lot No = 16512080A2
Cyl No = 8
Exp Date = 08/01/2014
County = 38 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = MOHL 8164.13.00 06/09
05/20/2014 14:51

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:52
02 Std. Gas	0.080	14:52
03 Room Air	0.000	14:53
04 Std. Gas	0.079	14:53
05 Room Air	0.000	14:54
06 Std. Gas	0.078	14:54
07 Room Air	0.000	14:55

Lot No = 16512080A2
Cyl No = 8
Exp Date = 08/01/2014
County = 38 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:
Calibration check 0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = MOHL 8164.13.00 06/09
05/20/2014 14:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:55
02 Std. Gas	0.080	14:56
03 Room Air	0.000	14:56
04 Std. Gas	0.079	14:57
05 Room Air	0.000	14:57
06 Std. Gas	0.078	14:57
07 Room Air	0.000	14:58

Lot No = 16512080A2
Cyl No = 8
Exp Date = 08/01/2014
County = 38 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC