



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59281 (1-2010)

Serial Number: 80-004953 Instrument Location: New Town (TAT) Police Dept.

Reason for Install/Repair:

Install After Lab Repair/Inspection

Other (Specify) _____

Check When Done:

- A. Surge Protector Installed/Properly Grounded.
- B. ~~Telephone Line Connected to Intoxilyzer® 8000.~~ N/A APX, Not Installed
- C. Breath Tube Heated.
- D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- E. Review/Enter Gas Setup (Level 1, Function S).
- F. Print Test (Level 1, Function P). (Attach Test Record.)
- G. ACA Test (Level 1, Function C). (Attach Test Record.)
- H. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- I. Repair and/or Maintenance Performed: _____

[Signature] 5-1-11
 Field Inspector Signature Steven R Fischer Date

Deb Kashur 31 May 2011
 Reviewed By EW9904 Date

Installation was not complete until 5-23-11.
 DK

80-004953
 P. 55 of 58
 PM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004953
Location = NEWT 8164.13.00 06/09
05/01/2011 18:40

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 709352
Standard Cyl #? 36
Standard Expiration? 12/23/2011
Oper No? 049904

Flow Cal. Date: 01/19/2010
Slope 691
Intercept -628068

IR Calibration Date: 06/08/2010
 3um 9um

0th Coef(*100): -10343 -18245
1st Coef(*100): 249276 131289
2nd Coef(*100): 4014 1439
H2O adj(mg/l*10k): 697 655

***** Printer Test End *****



Operator Signature
STEVEN R. FISCHER

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004953
Location = NEWT 8164.13.00 06/09
05/01/2011 18:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:43
02 Std. Gas	0.081	18:44
03 Room Air	0.000	18:44
04 Std. Gas	0.081	18:44
05 Room Air	0.000	18:45
06 Std. Gas	0.081	18:45
07 Room Air	0.000	18:46

Lot No = 709352
Cyl No = 36
Exp Date = 12/23/2011
County = 31 Oper No. = 049904



Operator Signature
STEVEN R. FISCHER

Remarks:

Form 106-I8000

80-004953
P. 57-58
JFM

