

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004934

Location: TOXL

- | | Check When Done |
|---|-----------------|
| A. Is the breath tube heated? | <u>✓</u> |
| B. Is the diagnostic check complete? | <u>✓</u> |
| C. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| D. Print test. (Attach test record.) | <u>✓</u> |
| E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3374</u> Lot No. <u>201501B</u> Rep AC <u>0.020</u> | <u>✓</u> |
| F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3841</u> Lot No. <u>201603A</u> Rep AC <u>0.300</u> | <u>✓</u> |
| G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN <u>DR5144</u> Lot No. <u>FCS4</u> Solution <u>0.16 AC + 0.05% acetone</u> | <u>✓</u> |
| H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) | <u>✓</u> |
| I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) <u>0.080 AC</u> | |

Lot No. 13314080A2 Cyl No. 53 Exp. Date 06/05/2016

Test 1 0.080 Test 4 0.078 Test 7 0.079

Test 2 0.080 Test 5 0.080 Test 8 0.079

Test 3 0.079 Test 6 0.080 Test 9 0.079

Average 0.079

J. Tank pressure, Level 3 Func D Sub G P: 952 psi Reg 950 psi

K. Remarks/Maintenance record: Annual inspection, replaced breath hose/simulator return o-rings, cleaned breath hose screen

20 May 16
Date

Deb Shanaver
Field Inspector's Signature

23 MAY 16 ✓
Date

Charles E. Eden
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004934
Location = TOXL 8164.13.00 06/09
05/20/2016 13:54

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 13314080A2
Standard Cyl #? 53
Standard Expiration? 06/05/2016
Oper No? 777777

Flow Cal. Date: 01/20/2010
Slope 698
Intercept -694859

IR Calibration Date: 04/08/2015
 3um 9um


0th Coef(*100): -27086 -37187
1st Coef(*100): 261158 135425
2nd Coef(*100): 1128 764
H2O adj(mg/l*10k): 493 494

***** Printer Test End *****



Operator Signature
DEB SHANAVER

Remarks:



Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004934
Location = TOXL 8164.13.00 06/09
05/20/2016 14:02

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:03
02 Std. Sol.	0.016	14:03
03 Room Air	0.000	14:04
04 Std. Sol.	0.017	14:05
05 Room Air	0.000	14:05
06 Std. Sol.	0.016	14:06
07 Room Air	0.000	14:06

08 Sim Temp = 34.0°C

Simul Ser No = DR3374

Std Sol No = 201501B

County = 08

Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

Low alcohol 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004934
Location = TOXL 8164.13.00 06/09
05/20/2016 14:08

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:09
02 Std. Sol.	0.297	14:09
03 Room Air	0.000	14:10
04 Std. Sol.	0.297	14:11
05 Room Air	0.000	14:11
06 Std. Sol.	0.297	14:12
07 Room Air	0.000	14:12

08 Sim Temp = 34.0°C

Simul Ser No = DR3841

Std Sol No = 201603A

County = 08

Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

High alcohol 0.300

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004934
Location = TOXL 8164.13.00 06/09
05/20/2016 13:58

Test	AC	Time
01 Room Air	0.000	13:58
02 *Subject Test	INT*	13:59
03 Room Air	0.000	13:59

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 13314080A2
Cyl No = 53
Expiration Date = 06/05/2016
County = 08 Oper No. = 777777



Operator Signature
DEB SHANAVER

Remarks:

Interferent detected 0.10AC + 0.05% acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004934
Location = TOXL 8164.13.00 06/09
05/20/2016 14:00

Test	AC	Time
01 Diagnostic	OK	14:00
02 Room Air	RFI*	14:00
03 Room Air	0.000	14:01

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 13314080A2
Cyl No = 53
Expiration Date = 06/05/2016
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB SHANAVER

Remarks:

RFI test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004934
Location = TOXL 8164.13.00 06/09
05/20/2016 13:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:45
02 Std. Gas	0.080	13:45
03 Room Air	0.000	13:46
04 Std. Gas	0.080	13:46
05 Room Air	0.000	13:47
06 Std. Gas	0.079	13:47
07 Room Air	0.000	13:48

Lot No = 13314080A2
Cyl No = 53
Exp Date = 06/05/2016
County = 08 Oper No. = 777777

Deb Shanaver
Operator Signature
DEB SHANAVER

Remarks:

Calibration check 0.080AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004934
Location = TOXL 8164.13.00 06/09
05/20/2016 13:49

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:49
02 Std. Gas	0.078	13:49
03 Room Air	0.000	13:50
04 Std. Gas	0.080	13:50
05 Room Air	0.000	13:51
06 Std. Gas	0.080	13:51
07 Room Air	0.000	13:52

Lot No = 13314080A2
Cyl No = 53
Exp Date = 06/05/2016
County = 08 Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

Calibration check 0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004934
Location = TOXL 8164.13.00 06/09
05/20/2016 13:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:55
02 Std. Gas	0.079	13:55
03 Room Air	0.000	13:56
04 Std. Gas	0.079	13:56
05 Room Air	0.000	13:56
06 Std. Gas	0.079	13:57
07 Room Air	0.000	13:57

Lot No = 13314080A2
Cyl No = 53
Exp Date = 06/05/2016
County = 08 Oper No. = 777777



Operator Signature
DEB SHANAVER

Remarks:

Calibration check 0.080AC

Form 106-I8000