



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59281 (1-2010)

Serial Number: <u>80-004943</u>	Instrument Location: <u>STAN</u>
Reason for Install/Repair: <input checked="" type="checkbox"/> Install After Lab Repair/Inspection <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- A. Surge Protector Installed/Properly Grounded.
- B. Telephone Line Connected to Intoxilyzer® 8000.
- C. Breath Tube Heated.
- D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- E. Review/Enter Gas Setup (Level 1, Function S).
- F. Print Test (Level 1, Function P). (Attach Test Record.)
- G. ACA Test (Level 1, Function C). (Attach Test Record.)
- H. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- I. Repair and/or Maintenance Performed: _____

[Signature]
Field Inspector Signature

05/24/10
Date

Deb Kashur
Reviewed By

26 May 2010
Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = STAN 8164.13.00 06/09
05/24/2010 20:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:29
02 Std. Gas	0.079	20:30
03 Room Air	0.000	20:30
04 Std. Gas	0.079	20:31
05 Room Air	0.000	20:31
06 Std. Gas	0.078	20:32
07 Room Air	0.000	20:32

Lot No = 709352
Cyl No = 34
Exp Date = 12/23/2011
County = 31 Oper No. = 100019


Operator Signature
DUANE L. MARMON

Remarks:

Form 106-I8000

80-004943
P. 45/46 KM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = STAN 8164.13.00 06/09
05/24/2010 20:33

Test	AC	Time
01 Room Air	RFI*	20:36
02 Room Air	0.000	20:36

*Invalid Test
Inhibited - RFI

Sub Name = TEST, SUBJECT A
Sub DOB = 01/01/1991
Sub Sex = Female Weight = 300
Test = OTH Cit = 12345
Dr. Lic. = ND/TES911234
Lot No = 709352
Cyl No = 34
Expiration Date = 12/23/2011
County = 31 Oper No. = 100019


Operator Signature
DUANE L. MARMON

Remarks:

Form 106-I8000

80-004943
P. 46/46 YM