



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name:	Shannon Liebald				
Location:	Mercer County Sheriff's Office				
Intoxilyzer® Serial Number:	80-003060				
Gas Lot Number:	659357	Gas Cylinder Number:	43	Expiration Date:	9-3-11
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed				
Instrument Test Results (Report to 3 Digits; Example, 0.081%)					
Test 1: 0.080 %					
Test 2: 0.080 %					
Test 3: 0.080 %					
(ATTACH TEST RECORD)					

Shannon Liebald
Chemical Test Operator Signature

5-26-10
Date

Deb Kashner
Reviewed By

28 May '10
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003060
Location = STAT 8164.13.00 06/09
05/26/2010 21:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	21:49
02 Std. Gas	0.080	21:50
03 Room Air	0.000	21:50
04 Std. Gas	0.080	21:50
05 Room Air	0.000	21:51
06 Std. Gas	0.080	21:51
07 Room Air	0.000	21:52

Lot No = 659357

Cyl No = 43

Exp Date = 09/03/2011

County = 29

Oper No. = 020904



Operator Signature
SHANNON P. LIEBOLD

Remarks:

Form 106-I8000

80-003060
P. 8/1/88 KM