

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004954

Location: TOXL

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5132 Lot No. --- Rep AC 0.021 ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5134 Lot No. --- Rep AC 0.294 ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3846 Lot No. --- Rep AC 0.10 AC + acetone ✓
- J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
- K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 15410080A1 Cyl No. 29 Exp. Date 07/01/2012
Test 1 0.079 Test 4 0.079 Test 7 0.079
Test 2 0.078 Test 5 0.079 Test 8 0.079
Test 3 0.079 Test 6 0.079 Test 9 0.080
Average 0.079

L. Tank pressure, Level 3 Func D Sub G P: 225 psi Reg 240 psi

M. Remarks/Maintenance record: Replaced cylinder regulator Annual inspection,
calibrated cyle regulator
26 May 2011 Deb Kashner
Date Field Inspector's Signature
S-26-11 [Signature]
Date Reviewed by

80-004954
P. 29 of 77
FM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = TOXL 8164.13.00 06/09
05/26/2011 11:32

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 15410080A1
Standard Cyl #? 29
Standard Expiration? 07/01/2012
Oper No? 777777

Flow Cal. Date: 01/19/2010
Slope 679
Intercept -671831

IR Calibration Date: 01/19/2010

	3um	9um
0th Coef(*100):	-34134	-33955
1st Coef(*100):	257873	132608
2nd Coef(*100):	1059	850
H2O adj(mg/l*10k):	762	516

***** Printer Test End *****

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Print test
Form 106-18000

80-004954
P. 70 of 77
AM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = TOXL 8164.13.00 06/09
05/26/2011 13:32

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:33
02 Std. Sol.	0.020	13:34
03 Room Air	0.000	13:34
04 Std. Sol.	0.019	13:35
05 Room Air	0.000	13:35
06 Std. Sol.	0.019	13:36
07 Room Air	0.000	13:37

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

low AC 0.021 AC

Form 106-I8000

80-004954
P. 71 of 77
FM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = TOXL 8164.13.00 06/09
05/26/2011 13:37

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:38
02 Std. Sol.	0.295	13:39
03 Room Air	0.000	13:40
04 Std. Sol.	0.293	13:40
05 Room Air	0.000	13:41
06 Std. Sol.	0.294	13:42
07 Room Air	0.000	13:42

08 Sim Temp = 34.0°C

Simul Ser No = 0
Std Sol No = 0
County = 08

Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.294 AC

Form 106-I8000

80-004954
p. 72 of 77
RM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = TOXL 8164.13.00 06/09
05/26/2011 13:44

Test	AC	Time
01 Room Air	0.000	13:45
02 *Subject Test	INT*	13:45
03 Room Air	0.000	13:46

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 15410080A1
Cyl No = 29
Expiration Date = 07/01/2012
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.10 AC + acetone

Form 106-I8000

80-004954
p. 73 of 77
JTH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = TOXL 8164.13.00 06/09
05/26/2011 13:46

Test	AC	Time
01 Diagnostic	OK	13:47
02 Room Air	RFI*	13:47
03 Room Air	0.000	13:47

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 15410080A1
Cyl No = 29
Expiration Date = 07/01/2012
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

RFI check

Form 106-I8000

80-004954
p. 74 of 77
VM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = TOXL 8164.13.00 06/09
05/26/2011 11:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:33
02 Std. Gas	0.079	11:34
03 Room Air	0.000	11:34
04 Std. Gas	0.078	11:35
05 Room Air	0.000	11:35
06 Std. Gas	0.079	11:35
07 Room Air	0.000	11:36

Lot No = 15410080A1
Cyl No = 29
Exp Date = 07/01/2012
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:
Calibration check 0.080 AC

Form 106-I8000

80-004954
P. 75 of 77
FR

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = TOXL 8164.13.00 06/09
05/26/2011 11:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:37
02 Std. Gas	0.079	11:37
03 Room Air	0.000	11:38
04 Std. Gas	0.079	11:38
05 Room Air	0.000	11:39
06 Std. Gas	0.079	11:39
07 Room Air	0.000	11:39

Lot No = 15410080A1
Cyl No = 29
Exp Date = 07/01/2012
County = 08 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-004954
p. 76 of 77
KM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = TOXL 8164.13.00 06/09
05/26/2011 11:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:41
02 Std. Gas	0.079	11:41
03 Room Air	0.000	11:41
04 Std. Gas	0.079	11:42
05 Room Air	0.000	11:42
06 Std. Gas	0.080	11:43
07 Room Air	0.000	11:43

Lot No = 15410080A1
Cyl No = 29
Exp Date = 07/01/2012
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080Ac

Form 106-I8000

80-004954
P. 77 of 77
KFM