

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-005941

Location: FTOT

(SPIRIT LAKE
FORT TOTTEN
62)

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DRS191 Lot No. 15160 Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3847 Lot No. 16060 Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DRS190 Lot No. ICS4 Solution 0.10% AL + 0.05% ACETONE ✓
 - H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 14214080AC6 Cyl No. 071 Exp. Date 07/05/2016

Test 1 <u>0.081</u>	Test 4 <u>0.080</u>	Test 7 <u>0.080</u>
Test 2 <u>0.079</u>	Test 5 <u>0.080</u>	Test 8 <u>0.080</u>
Test 3 <u>0.080</u>	Test 6 <u>0.080</u>	Test 9 <u>0.080</u>

Average 0.080

J. Tank pressure, Level 3 Func D Sub G P: 862 psi Reg ~850 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

5/25/2016
Date

[Signature]
Field Inspector's Signature

26 May 16
Date

[Signature]
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = FTOT 8164.13.00 06/09
05/25/2016 12:35

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 14214080A6
Standard Cyl #? 71
Standard Expiration? 07/05/2016
Oper No? 666666

Flow Cal. Date: 06/18/2013
Slope 695
Intercept -675557

IR Calibration Date: 06/18/2013
 3um 9um

0th Coef(*100): -20306 -25137
1st Coef(*100): 291687 138772
2nd Coef(*100): 2450 1058
H2O adj(mg/l*10k): 562 452

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks: PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = FTOT 8164.13.00 06/09
05/25/2016 12:45

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:45
02 Std. Sol.	0.016	12:46
03 Room Air	0.000	12:47
04 Std. Sol.	0.017	12:47
05 Room Air	0.000	12:48
06 Std. Sol.	0.017	12:49
07 Room Air	0.000	12:49


08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 15160

County = 62

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC

0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = FTOT 8164.13.00 06/09
05/25/2016 12:56

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:57
02 Std. Sol.	0.304	12:57
03 Room Air	0.000	12:58
04 Std. Sol.	0.304	12:59
05 Room Air	0.000	12:59
06 Std. Sol.	0.304	13:00
07 Room Air	0.000	13:01

08 Sim Temp = 34.0°C

Simul Ser No = DR3847

Std Sol No = 16060

County = 62

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = FTOT 8164.13.00 06/09
05/25/2016 12:41

Test	AC	Time
01 Room Air	0.000	12:42
02 *Subject Test	INT*	12:42
03 Room Air	0.000	12:43

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 14214080A6
Cyl No = 71
Expiration Date = 07/05/2016
County = 62 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent Check*
0.16 AC + 0.05% ACETONE
Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = FTOT 8164.13.00 06/09
05/25/2016 12:39

Test	AC	Time
01 Diagnostic	OK	12:39
02 Room Air	0.000	12:40
03 *Subject Test	RFI*	12:40
04 Room Air	0.000	12:40

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 14214080A6
Cyl No = 71
Expiration Date = 07/05/2016
County = 62 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = FTOT 8164.13.00 06/09
05/25/2016 13:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:05
02 Std. Gas	0.081	13:05
03 Room Air	0.000	13:05
04 Std. Gas	0.079	13:06
05 Room Air	0.000	13:06
06 Std. Gas	0.080	13:07
07 Room Air	0.000	13:07

Lot No = 14214080A6
Cyl No = 71
Exp Date = 07/05/2016
County = 62 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = FTOT 8164.13.00 06/09
05/25/2016 13:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:08
02 Std. Gas	0.080	13:08
03 Room Air	0.000	13:09
04 Std. Gas	0.080	13:09
05 Room Air	0.000	13:10
06 Std. Gas	0.080	13:10
07 Room Air	0.000	13:11

Lot No = 14214080A6
Cyl No = 71
Exp Date = 07/05/2016
County = 62 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005941
Location = FTOT 8164.13.00 06/09
05/25/2016 13:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:11
02 Std. Gas	0.080	13:12
03 Room Air	0.000	13:12
04 Std. Gas	0.080	13:13
05 Room Air	0.000	13:13
06 Std. Gas	0.080	13:13
07 Room Air	0.000	13:14

Lot No = 14214080A6
Cyl No = 71
Exp Date = 07/05/2016
County = 62 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC