

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-005954

Location: DEVL

(LAKE REGION L.E.C.)
36

Check When Done

- A. Is the breath tube heated? ✓
- B. Is the diagnostic check complete? ✓
- C. Is the time, date, and year correct (re-set if necessary)? ✓
- D. Print test. (Attach test record.) ✓
- E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5191 Lot No. 15160 Rep AC 0.020 ✓
- F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3847 Lot No. 16060 Rep AC 0.300 ✓
- G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR5190 Lot No. ICS4 Solution 0.10AC + 0.05% ACETONE ✓
- H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
- I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 13314080A3 Cyl No. 36 Exp. Date 06/05/2016

Test 1 <u>0.079</u>	Test 4 <u>0.079</u>	Test 7 <u>0.079</u>
Test 2 <u>0.078</u>	Test 5 <u>0.079</u>	Test 8 <u>0.079</u>
Test 3 <u>0.079</u>	Test 6 <u>0.078</u>	Test 9 <u>0.078</u>

Average 0.079

J. Tank pressure, Level 3 Func D Sub G P: 366 psi Reg ~375 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

5/25/2016
Date

Charles E. Edler
Field Inspector's Signature

26 May 16 ✓
Date

Dab Shanaver
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = DEVL 8164.13.00 06/09
05/25/2016 14:10

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:10
02 Std. Sol.	0.017	14:11
03 Room Air	0.000	14:12
04 Std. Sol.	0.018	14:12
05 Room Air	0.000	14:13
06 Std. Sol.	0.018	14:14
07 Room Air	0.000	14:14

08 Sim Temp = 34.0°C

Simul Ser No = DR5191

Std Sol No = 15160

County = 36

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

Low AC

0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = DEVL 8164.13.00 06/09
05/25/2016 14:05

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:05
02 Std. Sol.	0.293	14:06
03 Room Air	0.000	14:07
04 Std. Sol.	0.293	14:07
05 Room Air	0.000	14:08
06 Std. Sol.	0.293	14:09
07 Room Air	0.000	14:09

08 Sim Temp = 34.0°C

Simul Ser No = DR3847
Std Sol No = 16060
County = 36 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *HIGH AC*
0.300 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = DEVL 8164.13.00 06/09
05/25/2016 13:48

Test	AC	Time
01 Room Air	0.000	13:49
02 *Subject Test	INT*	13:49
03 Room Air	0.000	13:50

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 13314080A3
Cyl No = 36
Expiration Date = 06/05/2016
County = 36 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = DEVL 8164.13.00 06/09
05/25/2016 13:51

Test	AC	Time
01 Diagnostic	OK	13:52
02 Room Air	0.000	13:53
03 *Subject Test	RFI*	13:53
04 Room Air	0.000	13:54

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 13314080A3
Cyl No = 36
Expiration Date = 06/05/2016
County = 36 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = DEVL 8164.13.00 06/09
05/25/2016 13:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:56
02 Std. Gas	0.079	13:57
03 Room Air	0.000	13:57
04 Std. Gas	0.078	13:57
05 Room Air	0.000	13:58
06 Std. Gas	0.079	13:58
07 Room Air	0.000	13:59

Lot No = 13314080A3
Cyl No = 36
Exp Date = 06/05/2016
County = 36 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = DEVL 8164.13.00 06/09
05/25/2016 13:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:00
02 Std. Gas	0.079	14:00
03 Room Air	0.000	14:01
04 Std. Gas	0.079	14:01
05 Room Air	0.000	14:02
06 Std. Gas	0.078	14:02
07 Room Air	0.000	14:03


Lot No = 13314080A3

Cyl No = 36

Exp Date = 06/05/2016

County = 36

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005954
Location = DEVL 8164.13.00 06/09
05/25/2016 14:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:22
02 Std. Gas	0.079	14:22
03 Room Air	0.000	14:23
04 Std. Gas	0.079	14:23
05 Room Air	0.000	14:24
06 Std. Gas	0.078	14:24
07 Room Air	0.000	14:25

Lot No = 13314080A3
Cyl No = 36
Exp Date = 06/05/2016
County = 36 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC