

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004204

Location: TOXL

- | | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3455</u> Lot No. <u>12210</u> Rep AC <u>0.020</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3847</u> Lot No. <u>13020</u> Rep AC <u>0.300</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR3377</u> Lot No. <u>ICS#1</u> Rep AC <u>0.100 AC + 0.05% ACETONE</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |

K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. <u>11611080A2</u>	Cyl No. <u>143</u>	Exp. Date <u>06/01/2013</u>
Test 1 <u>0.079</u>	Test 4 <u>0.079</u>	Test 7 <u>0.079</u>
Test 2 <u>0.079</u>	Test 5 <u>0.079</u>	Test 8 <u>0.079</u>
Test 3 <u>0.079</u>	Test 6 <u>0.079</u>	Test 9 <u>0.079</u>
Average <u>0.079</u>		

L. Tank pressure, Level 3 Func D Sub G P: 388 psi Reg 400 psi

M. Remarks/Maintenance record: ANNUAL INSPECTION
REPLACED EXTERNAL BATTERY 5/22/13; REPLACED BREATH
TUBE O-RING

24 MAY 13
Date

Charles E. Eden
Field Inspector's Signature

28 May 13
Date

Deb Kashur
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = TOXL 8164.13.00 06/09
05/24/2013 13:23

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 11611080A2
Standard Cyl #? 143
Standard Expiration? 06/01/2013
Oper No? 666666

Flow Cal. Date: 07/13/2010
Slope 750
Intercept -815939

IR Calibration Date: 07/14/2011
3um 9um

0th Coef(*100): -9686 -4209
1st Coef(*100): 259274 128761
2nd Coef(*100): 3107 1691
H2O adj(mg/l*10k): 532 548

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = TOXL 8164.13.00 06/09
05/24/2013 13:30

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:30
02 Std. Sol.	0.022	13:31
03 Room Air	0.000	13:32
04 Std. Sol.	0.022	13:32
05 Room Air	0.000	13:33
06 Std. Sol.	0.021	13:34
07 Room Air	0.000	13:34

08 Sim Temp = 34.0°C

Simul Ser No = DR3455
Std Sol No = 12210
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *LOW AC*
 0.020 AC (Sim. Sol.)

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = TOXL 8164.13.00 06/09
05/24/2013 13:35

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:35
02 Std. Sol.	0.297	13:41
03 Room Air	0.000	13:41
04 Std. Sol.	0.298	13:42
05 Room Air	0.000	13:43
06 Std. Sol.	0.298	13:43
07 Room Air	0.000	13:44


08 Sim Temp = 34.0°C

Simul Ser No = DR3847

Std Sol No = 13020

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

LINEARITY TEST

0.300 AC (Sim. Sol.)

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = TOXL 8164.13.00 06/09
05/24/2013 13:58

Test	AC	Time
01 Room Air	0.000	13:59
02 *Subject Test	INT*	14:00
03 Room Air	0.000	14:00

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = QWERTYUIOPASDFG
Dr. Lic. = ND/DIS821456
Lot No = 11611080A2
Cyl No = 143
Expiration Date = 06/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT CHECK*
0.100 AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = TOXL 8164.13.00 06/09
05/24/2013 14:03

Test	AC	Time
01 Diagnostic	OK	14:04
02 Room Air	RFI*	14:04
03 Room Air	0.000	14:04

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = MNBVCXZLKJH0987
Dr. Lic. = ND/DIS821456
Lot No = 11611080A2
Cyl No = 143
Expiration Date = 06/01/2013
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *R.F.I. CHECK*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = TOXL 8164.13.00 06/09
05/24/2013 14:05

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:05
02 Std. Gas	0.079	14:06
03 Room Air	0.000	14:06
04 Std. Gas	0.079	14:07
05 Room Air	0.000	14:07
06 Std. Gas	0.079	14:07
07 Room Air	0.000	14:08

Lot No = 11611080A2
Cyl No = 143
Exp Date = 06/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC (E.G.S.)

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = TOXL 8164.13.00 06/09
05/24/2013 14:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:09
02 Std. Gas	0.079	14:09
03 Room Air	0.000	14:10
04 Std. Gas	0.079	14:10
05 Room Air	0.000	14:10
06 Std. Gas	0.079	14:11
07 Room Air	0.000	14:11

Lot No = 11611080A2
Cyl No = 143
Exp Date = 06/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080AC (E.G.S.)

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004204
Location = TOXL 8164.13.00 06/09
05/24/2013 14:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:12
02 Std. Gas	0.079	14:12
03 Room Air	0.000	14:13
04 Std. Gas	0.079	14:13
05 Room Air	0.000	14:14
06 Std. Gas	0.079	14:14
07 Room Air	0.000	14:15

Lot No = 11611080A2

Cyl No = 143

Exp Date = 06/01/2013

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
.0.080 AC (E.G.S.)

Form 106-I8000