

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004937

Location: TOXL

- | | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3455</u> Lot No. <u>12210</u> Rep AC <u>0.020</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3847</u> Lot No. <u>13020</u> Rep AC <u>0.300</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR377</u> Lot No. <u>ICS#1</u> Rep AC <u>0.100AC + 0.05% ACETONE</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. 11611080A2 Cyl No. 143 Exp. Date 06/01/2013
Test 1 0.080 Test 4 0.081 Test 7 0.081
Test 2 0.080 Test 5 0.080 Test 8 0.080
Test 3 0.080 Test 6 0.080 Test 9 0.081
Average 0.080

L. Tank pressure, Level 3 Func D Sub G P: 405 psi Reg 400 psi

M. Remarks/Maintenance record: ANNUAL INSPECTION
REPLACED EXTERNAL BATTERY 5/22/13; REPLACED BREATH TUBE O-RING

24 MAY 13
Date

Charles E. Eder
Field Inspector's Signature

28 May 13
Date

Deb Kashur
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004937
Location = TOXL 8164.13.00 06/09
05/24/2013 09:09

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 11611080A2
Standard Cyl #? 143
Standard Expiration? 06/01/2013
Oper No? 666666

Flow Cal. Date: 01/18/2010
Slope 730
Intercept -779991

IR Calibration Date: 05/24/2011
 3um 9um

0th Coef(*100): -12537 -15884
1st Coef(*100): 262613 132554
2nd Coef(*100): 1600 951
H2O adj(mg/l*10k): 529 537

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004937
Location = TOXL 8164.13.00 06/09
05/24/2013 09:17

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:17
02 Std. Sol.	0.021	09:18
03 Room Air	0.000	09:19
04 Std. Sol.	0.020	09:19
05 Room Air	0.000	09:20
06 Std. Sol.	0.020	09:21
07 Room Air	0.000	09:21

08 Sim Temp = 34.0°C

Simul Ser No = DR3455

Std Sol No = 11210

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low ϕ CEE AC*

0.020 AC (Sim. Sol.)

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004937
Location = TOXL 8164.13.00 06/09
05/24/2013 09:22

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:22
02 Std. Sol.	0.294	09:23
03 Room Air	0.000	09:24
04 Std. Sol.	0.295	09:25
05 Room Air	0.000	09:25
06 Std. Sol.	0.295	09:26
07 Room Air	0.000	09:26


08 Sim Temp = 34.0°C

Simul Ser No = DR3847

Std Sol No = 13020

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

LINEARITY TEST

0.300AC (Sim.Sol.)

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004937
Location = TOXL 8164.13.00 06/09
05/24/2013 09:34

Test	AC	Time
01 Room Air	0.000	09:35
02 *Subject Test	INT*	09:36
03 Room Air	0.000	09:36

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = QWERTYUIOPASDFG
Dr. Lic. = ND/DIS821456
Lot No = 11611080A2
Cyl No = 143
Expiration Date = 06/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

INTERFERENT CHECK
0.106AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004937
Location = TOXL 8164.13.00 06/09
05/24/2013 10:32

Test	AC	Time
01 Diagnostic	OK	10:33
02 Room Air	RFI*	10:33
03 Room Air	0.000	10:33

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = MNBVCXZLKJH0987
Dr. Lic. = ND/DIS821456
Lot No = 11611080A2
Cyl No = 143
Expiration Date = 06/01/2013
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RF.I. CHECK*


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004937
Location = TOXL 8164.13.00 06/09
05/24/2013 10:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:36
02 Std. Gas	0.080	10:36
03 Room Air	0.000	10:37
04 Std. Gas	0.080	10:37
05 Room Air	0.000	10:37
06 Std. Gas	0.080	10:38
07 Room Air	0.000	10:38

Lot No = 11611080A2
Cyl No = 143
Exp Date = 06/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC (E.G.S.)


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004937
Location = TOXL 8164.13.00 06/09
05/24/2013 10:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:39
02 Std. Gas	0.081	10:39
03 Room Air	0.000	10:40
04 Std. Gas	0.080	10:40
05 Room Air	0.000	10:41
06 Std. Gas	0.080	10:41
07 Room Air	0.000	10:42

Lot No = 11611080A2
Cyl No = 143
Exp Date = 06/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC (E.G.S.)

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004937
Location = TOXL 8164.13.00 06/09
05/24/2013 10:42

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:43
02 Std. Gas	0.081	10:43
03 Room Air	0.000	10:44
04 Std. Gas	0.080	10:44
05 Room Air	0.000	10:44
06 Std. Gas	0.081	10:45
07 Room Air	0.000	10:45

Lot No = 11611080A2
Cyl No = 143
Exp Date = 06/01/2013
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC (E.G.S.)

Form 106-I8000