

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004943

Location: WATF (MCKENZIE COUNTY S.O.)

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5115 Lot No. 12111 Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5143 Lot No. 13080 Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3375 Lot No. ICS #2 Rep AC 0.100 AC + 0.05% ACETONE ✓
 - H. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 16512080A2 Cyl No. 062 Exp. Date 08/01/2014

Test 1 <u>0.079</u>	Test 4 <u>0.079</u>	Test 7 <u>0.078</u>
Test 2 <u>0.079</u>	Test 5 <u>0.078</u>	Test 8 <u>0.079</u>
Test 3 <u>0.079</u>	Test 6 <u>0.079</u>	Test 9 <u>0.079</u>

Average 0.079

J. Tank pressure, Level 3 Func D Sub G P: 665 psi Reg 650 psi

K. Remarks/Maintenance record: REPLACED SIMULATOR RETURN O-RING,
ANNUAL INSPECTION

29 MAY 14
Date

Charles E. Edin
Field Inspector's Signature

29 May 14
Date

Deb Kashner
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = WATF 8164.13.00 06/09
05/29/2014 10:31

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 16512080A2
Standard Cyl #? 62
Standard Expiration? 08/01/2014
Oper No? 666666

Flow Cal. Date: 01/18/2010
Slope 693
Intercept -665387

IR Calibration Date: 05/07/2013
 3um 9um

0th Coef(*100): -10324 -14602
1st Coef(*100): 231940 125849
2nd Coef(*100): 6794 2184
H2O adj(mg/l*10k): 586 616

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = WATF 8164.13.00 06/09
05/29/2014 10:49

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:50
02 Std. Sol.	0.017	10:51
03 Room Air	0.000	10:51
04 Std. Sol.	0.018	10:52
05 Room Air	0.000	10:53
06 Std. Sol.	0.018	10:53
07 Room Air	0.000	10:54


08 Sim Temp = 34.0°C

Simul Ser No = DR5115

Std Sol No = 12111

County = 27

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = WATF 8164.13.00 06/09
05/29/2014 11:00

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:01
02 Std. Sol.	0.310	11:02
03 Room Air	0.000	11:02
04 Std. Sol.	0.312	11:03
05 Room Air	0.000	11:04
06 Std. Sol.	0.314	11:04
07 Room Air	0.000	11:05

08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 13080

County = 27

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC

0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = WATF 8164.13.00 06/09
05/29/2014 10:35

Test	AC	Time
01 Room Air	0.000	10:36
02 *Subject Test	INT*	10:37
03 Room Air	0.000	10:37

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 16512080A2
Cyl No = 62
Expiration Date = 08/01/2014
County = 27 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

*INTERFERENT
CHECK
0.100AC + 0.05% ACETONE*
Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = WATF 8164.13.00 06/09
05/29/2014 10:32

Test	AC	Time
01 Diagnostic	OK	10:34
02 Room Air	RFI*	10:34
03 Room Air	0.000	10:34

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 16512080A2
Cyl No = 62
Expiration Date = 08/01/2014
County = 27 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = WATF 8164.13.00 06/09
05/29/2014 10:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:41
02 Std. Gas	0.079	10:41
03 Room Air	0.000	10:42
04 Std. Gas	0.079	10:42
05 Room Air	0.000	10:42
06 Std. Gas	0.079	10:43
07 Room Air	0.000	10:43

Lot No = 16512080A2
Cyl No = 62
Exp Date = 08/01/2014
County = 27 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *BOSE*
CALIBRATION
CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = WATF 8164.13.00 06/09
05/29/2014 10:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:45
02 Std. Gas	0.079	10:45
03 Room Air	0.000	10:46
04 Std. Gas	0.078	10:46
05 Room Air	0.000	10:47
06 Std. Gas	0.079	10:47
07 Room Air	0.000	10:47

Lot No = 16512080A2
Cyl No = 62
Exp Date = 08/01/2014
County = 27 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION
CHECK
0.080AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004943
Location = WATF 8164.13.00 06/09
05/29/2014 11:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:08
02 Std. Gas	0.078	11:08
03 Room Air	0.000	11:09
04 Std. Gas	0.079	11:09
05 Room Air	0.000	11:10
06 Std. Gas	0.079	11:10
07 Room Air	0.000	11:10

Lot No = 16512080A2
Cyl No = 62
Exp Date = 08/01/2014
County = 27 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION
CHECK
0.080 AC*

Form 106-I8000