



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Chad L. Harmon</u>		
Location: <u>FYPD</u>		
Intoxilyzer® Serial Number: <u>80-004954</u>		
Gas Lot Number: <u>709352</u>	Gas Cylinder Number: <u>26</u>	Expiration Date: <u>12-23-11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>079</u> %		
Test 2: 0. <u>079</u> %		
Test 3: 0. <u>079</u> %		
(ATTACH TEST RECORD)		

Chad L. Harmon
Chemical Test Operator Signature

05-31-10
Date

Deb Kashner
Reviewed By

03 June 2010
Date

80-004954
p. 47 of 48
sch

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = FYPD 8164.13.00 06/09
05/31/2010 03:09

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	03:10
02 Std. Gas	0.079	03:10
03 Room Air	0.000	03:11
04 Std. Gas	0.079	03:11
05 Room Air	0.000	03:12
06 Std. Gas	0.079	03:12
07 Room Air	0.000	03:13

Lot No = 709352
Cyl No = 26
Exp Date = 12/23/2011
County = 61 Oper No. = 010606



Operator Signature
CHAD L. HARMON

Remarks:

Form 106-I8000

80-004954
p. 48 of 48
KM