

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004946

Location: LINT

- | | Check When Done |
|---|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3845</u> Lot No. <u>13070</u> Rep AC <u>0.020</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3841</u> Lot No. <u>13020</u> Rep AC <u>0.300</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR5191</u> Lot No. <u>ICS #1</u> Rep AC <u>0.100 AC + acetone</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. <u>23411080A1</u>	Cyl No. <u>9</u>	Exp. Date <u>10/01/2013</u>
Test 1 <u>0.080</u>	Test 4 <u>0.080</u>	Test 7 <u>0.080</u>
Test 2 <u>0.080</u>	Test 5 <u>0.079</u>	Test 8 <u>0.080</u>
Test 3 <u>0.080</u>	Test 6 <u>0.080</u>	Test 9 <u>0.080</u>
Average <u>0.080</u>		

L. Tank pressure, Level 3 Func D Sub G P: 981 psi Reg 990 psi

M. Remarks/Maintenance record: Annual inspection, replaced external battery & gas
0-rings on breath hose & simulator return.
Deb Kashner
30 May 13 Date
31 MAY 13 Date
Field Inspector's Signature Charles E. Eden
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = LINT 8164.13.00 06/09
05/30/2013 13:59

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 23411080A1
Standard Cyl #? 9
Standard Expiration? 10/01/2013
Oper No? 777777

Flow Cal. Date: 01/18/2010
Slope 717
Intercept -728279

IR Calibration Date: 05/19/2011
 3um 9um

0th Coef(*100): -2231 -4869
1st Coef(*100): 282398 143476
2nd Coef(*100): 146 780
H2O adj(mg/l*10k): 303 375

***** Printer Test End *****



Operator Signature
DEB KASHUR

Remarks:

Print test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = LINT 8164.13.00 06/09
05/30/2013 14:06

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:06
02 Std. Sol.	0.022	14:07
03 Room Air	0.000	14:08
04 Std. Sol.	0.022	14:08
05 Room Air	0.000	14:09
06 Std. Sol.	0.022	14:10
07 Room Air	0.000	14:10

08 Sim Temp = 34.0°C

Simul Ser No = DR3845

Std Sol No = 13070

County = 15

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Low AC 0.020 AC

Form 106-18000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = LINT 8164.13.00 06/09
05/30/2013 14:12

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:12
02 Std. Sol.	0.302	14:13
03 Room Air	0.000	14:14
04 Std. Sol.	0.298	14:14
05 Room Air	0.000	14:15
06 Std. Sol.	0.295	14:16
07 Room Air	0.000	14:16

08 Sim Temp = 34.0°C

Simul Ser No = DR3841

Std Sol No = 13020

County = 15

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = LINT 8164.13.00 06/09
05/30/2013 14:21

Test	AC	Time
01 Room Air	0.000	14:21
02 *Subject Test	INT*	14:22
03 Room Air	0.000	14:22

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 9
Expiration Date = 10/01/2013
County = 15 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

*Interferent test 0.100 AC
+ acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = LINT 8164.13.00 06/09
05/30/2013 14:23

Test	AC	Time
01 Diagnostic	OK	14:24
02 Room Air	RFI*	14:24
03 Room Air	0.000	14:25

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 23411080A1
Cyl No = 9
Expiration Date = 10/01/2013
County = 15 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB KASHUR

Remarks:

RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = LINT 8164.13.00 06/09
05/30/2013 13:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:00
02 Std. Gas	0.080	14:00
03 Room Air	0.000	14:01
04 Std. Gas	0.080	14:01
05 Room Air	0.000	14:02
06 Std. Gas	0.080	14:02
07 Room Air	0.000	14:03

Lot No = 23411080A1
Cyl No = 9
Exp Date = 10/01/2013
County = 15 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = LINT 8164.13.00 06/09
05/30/2013 14:25

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:26
02 Std. Gas	0.080	14:26
03 Room Air	0.000	14:27
04 Std. Gas	0.079	14:27
05 Room Air	0.000	14:28
06 Std. Gas	0.080	14:28
07 Room Air	0.000	14:29

Lot No = 23411080A1
Cyl No = 9
Exp Date = 10/01/2013
County = 15 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004946
Location = LINT 8164.13.00 06/09
05/30/2013 14:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:30
02 Std. Gas	0.080	14:30
03 Room Air	0.000	14:31
04 Std. Gas	0.080	14:31
05 Room Air	0.000	14:32
06 Std. Gas	0.080	14:32
07 Room Air	0.000	14:33

Lot No = 23411080A1
Cyl No = 9
Exp Date = 10/01/2013
County = 15 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC