

INTOXILYZER 8000 INSPECTION

Serial No.: 80-002667

Location: TOXL

- |  | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes?   | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?  | <u>✓</u>        |
| C. Is the breath tube heated?  | <u>✓</u>        |
| D. Is the diagnostic check complete?   | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?  | <u>✓</u>        |
| F. Print test. (Attach test record.)   | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5144</u> Lot No. <u>    </u> Rep AC <u>0.021</u>   | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5114</u> Lot No. <u>    </u> Rep AC <u>0.298</u>                                | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR5186</u> Lot No. <u>    </u> Rep AC <u>0.100</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)  | <u>✓</u>        |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)  |                 |

|                       |                     |                              |
|-----------------------|---------------------|------------------------------|
| Lot No. <u>709352</u> | Cyl No. <u>44</u>   | Exp. Date <u>23 Dec 2011</u> |
| Test 1 <u>0.079</u>   | Test 4 <u>0.079</u> | Test 7 <u>0.080</u>          |
| Test 2 <u>0.080</u>   | Test 5 <u>0.079</u> | Test 8 <u>0.080</u>          |
| Test 3 <u>0.080</u>   | Test 6 <u>0.079</u> | Test 9 <u>0.079</u>          |
| Average <u>0.079</u>  |                     |                              |

L. Tank pressure, Level 3 Func D Sub G P: 203 psi Reg 200 psi

M. Remarks/Maintenance record: Annual inspection

01 June '10  
Date

Deb Kashner  
Field Inspector's Signature

14 Jun 10  
Date

Mary Plaron  
Reviewed by

80-002667  
P. 62 of 70  
sam

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-002667  
Location = TOXL                    8164.13.00 06/09  
06/01/2010                            10:56

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^&\*()\_+?

Current Instrument Setup


Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            822401I  
Standard Cyl #?            3  
Standard Expiration?        08/20/2010  
Oper No?                    777777

Flow Cal. Date:            09/10/2007  
Slope                        693  
Intercept                   -570251

IR Calibration Date:        09/10/2007  
                                  3um                            9um

-----  
0th Coef(\*100):            -21400                        -13991  
1st Coef(\*100):            268177                        135590  
2nd Coef(\*100):            1202                            658  
H2O adj(mg/l\*10k):        369                            165

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
Operator Signature  
DEB KASHUR

Remarks:

*Print Test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002667  
Location = TOXL      8164.13.00 06/09  
06/01/2010      11:18

WET CAL CHECK

| Test         | AC    | Time  |
|--------------|-------|-------|
| 01 Room Air  | 0.000 | 11:18 |
| 02 Std. Sol. | 0.019 | 11:19 |
| 03 Room Air  | 0.000 | 11:20 |
| 04 Std. Sol. | 0.019 | 11:20 |
| 05 Room Air  | 0.000 | 11:21 |
| 06 Std. Sol. | 0.020 | 11:22 |
| 07 Room Air  | 0.000 | 11:22 |

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

Low AC 0.021AC

Form 106-18000

80-002467  
P. 64/70  
JK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002667  
Location = TOXL      8164.13.00 06/09  
06/01/2010      11:23

WET CAL CHECK

| Test         | AC    | Time  |
|--------------|-------|-------|
| 01 Room Air  | 0.000 | 11:24 |
| 02 Std. Sol. | 0.289 | 11:25 |
| 03 Room Air  | 0.000 | 11:25 |
| 04 Std. Sol. | 0.291 | 11:26 |
| 05 Room Air  | 0.000 | 11:27 |
| 06 Std. Sol. | 0.291 | 11:27 |
| 07 Room Air  | 0.000 | 11:28 |

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

*Deb Kasher*

Operator Signature

DEB KASHUR

Remarks:

*Linearity test 0.298 AC*

Form 106-18000

80-002667  
P. 65 of 70  
PAL

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002667  
Location = TOXL      8164.13.00 06/09  
06/01/2010      13:59

| Test             | AC    | Time  |
|------------------|-------|-------|
| 01 Room Air      | 0.000 | 14:00 |
| 02 *Subject Test | INT*  | 14:00 |
| 03 Room Air      | 0.000 | 14:00 |

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 709352  
Cyl No = 44  
Expiration Date = 12/23/2011  
County = 08      Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Interferent check      0.100 AC + acetone*

Form 106-I8000

80-002667  
p. 66 of 70  
JAN

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002667  
Location = TOXL      8164.13.00 06/09  
06/01/2010      14:02

| Test          | AC    | Time  |
|---------------|-------|-------|
| 01 Diagnostic | OK    | 14:03 |
| 02 Room Air   | RFI*  | 14:03 |
| 03 Room Air   | 0.000 | 14:03 |

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 709352  
Cyl No = 44  
Expiration Date = 12/23/2011  
County = 08      Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*RFI check*

Form 106-I8000

80-002667  
p. 6 of 70  
yfm

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002667  
Location = TOXL      8164.13.00 06/09  
06/01/2010      14:07

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 14:08 |
| 02 Std. Gas | 0.079 | 14:08 |
| 03 Room Air | 0.000 | 14:08 |
| 04 Std. Gas | 0.080 | 14:09 |
| 05 Room Air | 0.000 | 14:09 |
| 06 Std. Gas | 0.080 | 14:10 |
| 07 Room Air | 0.000 | 14:10 |

Lot No = 709352  
Cyl No = 44  
Exp Date = 12/23/2011  
County = 08      Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000

80-002667  
P. 68 of 70  
JK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002667  
Location = TOXL      8164.13.00 06/09  
06/01/2010      14:11

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 14:11 |
| 02 Std. Gas | 0.079 | 14:12 |
| 03 Room Air | 0.000 | 14:12 |
| 04 Std. Gas | 0.079 | 14:12 |
| 05 Room Air | 0.000 | 14:13 |
| 06 Std. Gas | 0.079 | 14:13 |
| 07 Room Air | 0.000 | 14:14 |

Lot No = 709352  
Cyl No = 44  
Exp Date = 12/23/2011  
County = 08      Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000

*80-002667  
P. 6/14/10  
JAN*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002667  
Location = TOXL      8164.13.00 06/09  
06/01/2010      14:21

DRY CAL CHECK

| Test        | AC    | Time  |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 14:21 |
| 02 Std. Gas | 0.080 | 14:21 |
| 03 Room Air | 0.000 | 14:22 |
| 04 Std. Gas | 0.080 | 14:22 |
| 05 Room Air | 0.000 | 14:23 |
| 06 Std. Gas | 0.079 | 14:23 |
| 07 Room Air | 0.000 | 14:24 |

Lot No = 709352  
Cyl No = 44  
Exp Date = 12/23/2011  
County = 08      Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Calibration check 0.080 AC*

Form 106-I8000

80-002667  
P. 70 of 70  
KM