



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name:	<i>David M Wallace</i>		
Location:	<i>Dickinson, I. E.</i>		
Intoxilyzer® Serial Number:	<i>80 - 003066</i>		
Gas Lot Number:	Gas Cylinder Number:	Expiration Date:	
<i>659358</i>	<i>23</i>	<i>9-3-11</i>	
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)			
Test 1:	<i>0.082</i>	%	
Test 2:	<i>0.082</i>	%	
Test 3:	<i>0.082</i>	%	
(ATTACH TEST RECORD)			

*DM Wallace*  
Chemical Test Operator Signature

*6-1-2010*  
Date

*Deb Kashur*  
Reviewed By

*03 June 2010*  
Date

*80-003066*  
*p. 95 of 96*  
*SM*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003066  
Location = DKPD      8164.13.00 06/09  
06/01/2010      08:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:44
02 Std. Gas	0.082	08:45
03 Room Air	0.000	08:45
04 Std. Gas	0.082	08:46
05 Room Air	0.000	08:46
06 Std. Gas	0.082	08:47
07 Room Air	0.000	08:47

Lot No = 659358  
Cyl No = 23  
Exp Date = 09/03/2011  
County = 45      Oper No. = 068332

  
Operator Signature  
DAVID M. WALLACE

Remarks:

Form 106-I8000

80-003066  
P. 96 of 96  
YMW