



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59281 (1-2010)

Serial Number: 80-004942 Instrument Location: LaMoure Co. Courthouse


Reason for Install/Repair:

Install After Lab Repair/Inspection

Other (Specify) _____

Check When Done:

- A. Surge Protector Installed/Properly Grounded.
- B. Telephone Line Connected to Intoxilyzer® 8000.
- C. Breath Tube Heated.
- D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- E. Review/Enter Gas Setup (Level 1, Function S).
- F. Print Test (Level 1, Function P). (Attach Test Record.)
- G. ACA Test (Level 1, Function C). (Attach Test Record.)
- H. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- I. Repair and/or Maintenance Performed: _____


 Field Inspector Signature

6-2-10
 Date

Deb Kashur
 Reviewed By

17 June 2010
 Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = LAMO 8164.13.00 06/09
06/02/2010 16:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:40
02 Std. Gas	0.079	16:40
03 Room Air	0.000	16:41
04 Std. Gas	0.078	16:41
05 Room Air	0.000	16:42
06 Std. Gas	0.078	16:42
07 Room Air	0.000	16:43

Lot No = 709352
Cyl No = 14
Exp Date = 12/23/2011
County = 23 Oper No. = 020828



Operator Signature
ADAM R. KRABBENHOFT

Remarks:

ACA TEST

Form 106-18000

80-004942
P. 45 of 46
YFM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = LAMO 8164.13.00 06/09
06/02/2010 16:45

Test	AC	Time
01 Diagnostic	OK	16:47
02 Room Air	0.000	16:47
03 *Subject Test	RFI*	16:48
04 Room Air	0.000	16:48

*Invalid Test
Inhibited - RFI

Sub Name = TEST, SUBJECT Q
Sub DOB = 01/01/1987
Sub Sex = Male Weight = 100
Test = DUI Cit = 0123456
Dr. Lic. = ND/TES870001
Lot No = 709352
Cyl No = 14
Expiration Date = 12/23/2011
County = 23 Oper No. = 020828

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ADAM R. KRABBENHOFT

Remarks:

RFI test

Form 106-18000

*80-004942
P. 46 of 46
AK*