



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name:	Adam Krabbenhoft				
Location:	LaMoore County Courthouse				
Intoxilyzer® Serial Number:	80-004942				
Gas Lot Number:	709352	Gas Cylinder Number:	14	Expiration Date:	12-23-11
<input checked="" type="checkbox"/> New Cylinder Installed	<input type="checkbox"/> Cylinder Not Changed				
Instrument Test Results (Report to 3 Digits; Example, 0.081%)					
Test 1: 0.078%					
Test 2: 0.079%					
Test 3: 0.080%					
(ATTACH TEST RECORD)					



Chemical Test Operator Signature

6-2-10

Date

Deb Kashur

Reviewed By

17 June 2010

Date

80-004942
P. 41 of 42
SKM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004942
Location = LAMO 8164.13.00 06/09
06/02/2010 16:25

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:26
02 Std. Gas	0.078	16:27
03 Room Air	0.000	16:27
04 Std. Gas	0.079	16:27
05 Room Air	0.000	16:28
06 Std. Gas	0.080	16:28
07 Room Air	0.000	16:29

Lot No = 709352
Cyl No = 14
Exp Date = 12/23/2011
County = 23 Oper No. = 020828



Operator Signature
ADAM R. KRABBENHOFT

Remarks: *ACA Test*

Form 106-I8000

*80-004942
P. 42 of 42
KRM*