



OFFICE OF ATTORNEY GENERAL
 Crime Laboratory Division
 2641 East Main Avenue
 P.O. Box 937
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Berry Vannotta

Location: Mohl

Intoxilyzer Serial No.: 80-004198

Lot Number: 659358

Cylinder Number: 27

Expiration Date: 9-3-11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.677 %

Test 2: 0.678 %

Test 3: 0.677 %

(ATTACH TEST RECORD)

Berry Vannotta
 Chemical Test Operator Signature

 Date Cylinder Changed

Deb Kashur 10 June 2010
 Reviewed By

6-3-10
 Date

Form 105-G
 Toxicology Laboratory
 Crime Laboratory Division

80-004198
 P.93/94 FM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = MOHL 8164.13.00 06/09
06/03/2010 21:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	21:59
02 Std. Gas	0.077	21:59
03 Room Air	0.000	22:00
04 Std. Gas	0.078	22:00
05 Room Air	0.000	22:00
06 Std. Gas	0.077	22:01
07 Room Air	0.000	22:01

Lot No = 659358
Cyl No = 27
Exp Date = 09/03/2011
County = 38 Oper No. = 039909


Operator Signature
BARRY A. VANNATTA

Remarks:

Form 106-18000

80-004198
P. 94/94 KM