



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name:	ZLOTKOWSKI, GERARD R				
Location:	ASHL				
Intoxilyzer® Serial Number:	80-004945				
Gas Lot Number:	709352	Gas Cylinder Number:	32	Expiration Date:	12/23/11
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed				
Instrument Test Results (Report to 3 Digits; Example, 0.081%)					
Test 1: 0. <u>079</u> %					
Test 2: 0. <u>079</u> %					
Test 3: 0. <u>079</u> %					
(ATTACH TEST RECORD)					



Chemical Test Operator Signature

6/3/2010

Date

Deb Kashner

Reviewed By

01 July 2010

Date

80-004945
p. 47-48
FK


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004945
Location = ASHL 8164.13.00 06/09
06/03/2010 15:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:28
02 Std. Gas	0.079	15:28
03 Room Air	0.000	15:28
04 Std. Gas	0.079	15:29
05 Room Air	0.000	15:29
06 Std. Gas	0.079	15:30
07 Room Air	0.000	15:30

Lot No = 709352
Cyl No = 32
Exp Date = 12/23/2011
County = 26 Oper No. = 100631



Operator Signature
GERARD R. ZLOTKOWSKI

Remarks:

Form 106-I8000

80-004945
P. 48 of 48
RM