

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004205

Location: Cass Co Jail

- | | Check When Done |
|---|-----------------|
| A. Is the warm-up time less than 20 minutes? | <u>✓</u> |
| B. Is a three-pronged grounded outlet used? | <u>✓</u> |
| C. Is the breath tube heated? | <u>✓</u> |
| D. Is the diagnostic check complete? | <u>✓</u> |
| E. Is the time, date, and year correct (re-set if necessary)? | <u>✓</u> |
| F. Print test. (Attach test record.) | <u>✓</u> |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN <u>DR5132</u> Lot No. <u>—</u> Rep AC <u>0.021 AC</u> | <u>✓</u> |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN <u>DR3375</u> Lot No. <u>—</u> Rep AC <u>0.292 AC</u> | <u>✓</u> |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN <u>DR3454</u> Lot No. <u>—</u> Rep AC <u>0.10 AC + 0.05% Acetone</u> | <u>✓</u> |
| J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) | <u>✓</u> |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) | |

Lot No. <u>709352</u>	Cyl No. <u>14</u>	Exp. Date <u>12/23/10</u>
Test 1 <u>0.080</u>	Test 4 <u>0.081</u> 0.080 MKP	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.080</u>	Test 8 <u>0.081</u>
Test 3 <u>0.080</u>	Test 6 <u>0.080</u>	Test 9 <u>0.081</u>
Average <u>0.081 AC</u>		

L. Tank pressure, Level 3 Func D Sub G P: 952 psi Reg 950 psi

M. Remarks/Maintenance record: Annual Inspection; Forms Load; New Gas Cylinder

3 June 2010
Date

Mary Plann
Field Inspector's Signature

14 June 2010
Date

Deb Kashur
Reviewed by

80-004205
P. 7/1/10

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = CASS 8164.13.00 06/09
06/03/2010 08:20

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Wet
Standard Value? 0.000
Standard Solution #? 00000
Sim Serial #? 00000000
Oper No? 107501

Flow Cal. Date: 06/04/2009
Slope 704
Intercept -610062

IR Calibration Date: 06/03/2009

	3um	9um
0th Coef(*100):	-16944	-29673
1st Coef(*100):	265939	132317
2nd Coef(*100):	1696	900
H2O adj(mg/l*10k):	490	583

***** Printer Test End *****



Operator Signature
MARGY PEARSON

Remarks:

Print

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = CASS 8164.13.00 06/09
06/03/2010 08:38

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:38
02 Std. Sol.	0.019	08:39
03 Room Air	0.000	08:40
04 Std. Sol.	0.018	08:41
05 Room Air	0.000	08:41
06 Std. Sol.	0.018	08:42
07 Room Air	0.000	08:42

08 Sim Temp = 34.0°C

Simul Ser No = 000000
Std Sol No = 00000
County = 09 Oper No. = 107501

Margy Pearson

Operator Signature
MARGY PEARSON

Remarks:

Low AC 0.021 AC

Form 106-I8000

*80-004205
P. 74 of 80
KJH*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = CASS 8164.13.00 06/09
06/03/2010 08:22

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:22
02 Std. Sol.	0.293	08:23
03 Room Air	0.000	08:24
04 Std. Sol.	0.292	08:24
05 Room Air	0.000	08:25
06 Std. Sol.	0.292	08:26
07 Room Air	0.000	08:26

08 Sim Temp = 34.0°C

Simul Ser No = 00000000

Std Sol No = 00000

County = 09

Oper No. = 107501

Margy Pearson

Operator Signature
MARGY PEARSON

Remarks:

Linearity 0.292 AC

Form 106-I8000

80-004205
P. 75 of 80
AM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = CASS 8164.13.00 06/09
06/03/2010 08:34

Test	AC	Time
01 Room Air	0.000	08:35
02 *Subject Test	INT*	08:36
03 Room Air	0.000	08:36

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 11/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS821456
Lot No = 709352
Cyl No = 14
Expiration Date = 12/23/2011
County = 09 Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

Interferent Alc + Acetone

Form 106-I8000

80-004205
P. 76 of 80
EM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = CASS 8164.13.00 06/09
06/03/2010 08:46

Test	AC	Time
01 Diagnostic	OK	08:46
02 Room Air	RFI*	08:47
03 Room Air	0.000	08:47

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 11/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS821456
Lot No = 709352
Cyl No = 14
Expiration Date = 12/23/2011
County = 09 Oper No. = 107501

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
MARGY PEARSON

Remarks:

RFI

Form 106-I8000

80-004205
P. 77 of 80
JRM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = CASS 8164.13.00 06/09
06/03/2010 08:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:51
02 Std. Gas	0.080	08:51
03 Room Air	0.000	08:52
04 Std. Gas	0.081	08:52
05 Room Air	0.000	08:53
06 Std. Gas	0.080	08:53
07 Room Air	0.000	08:54

Lot No = 709352
Cyl No = 14
Exp Date = 12/23/2011
County = 08 Oper No. = 107501

Margy Pearson

Operator Signature
MARGY PEARSON

Remarks:

Cal Check 0.080 AC

Form 106-I8000

80-004205
P. 78 of 80
JH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = CASS 8164.13.00 06/09
06/03/2010 08:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:54
02 Std. Gas	0.081	08:55
03 Room Air	0.000	08:55
04 Std. Gas	0.080	08:56
05 Room Air	0.000	08:56
06 Std. Gas	0.080	08:56
07 Room Air	0.000	08:57

Lot No = 709352
Cyl No = 14
Exp Date = 12/23/2011
County = 09 Oper No. = 107501

Margy Pearson

Operator Signature
MARGY PEARSON

Remarks:

Cal Check - AEA 0.080 AC

Form 106-I8000

80-004205
P. 79-480
AM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004205
Location = CASS 8164.13.00 06/09
06/03/2010 08:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:58
02 Std. Gas	0.081	08:58
03 Room Air	0.000	08:59
04 Std. Gas	0.081	08:59
05 Room Air	0.000	08:59
06 Std. Gas	0.081	09:00
07 Room Air	0.000	09:00

Lot No = 709352
Cyl No = 14
Exp Date = 12/23/2011
County = 09 Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

Cal Check 0.080 AC

Form 106-I8000

*80-004205
P. 80 of 80
RM*