



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name:	<i>Chad L. Harmon</i>		
Location:	<i>Fort Yates</i>		
Intoxilyzer® Serial Number:	<i>80-004954</i>		
Gas Lot Number:	<i>26</i> ⁷⁰⁹³⁵² <i>DK</i>	Gas Cylinder Number:	<i>709352</i> ²⁶ <i>DK</i> Expiration Date: <i>12-23-11</i>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed <i>Moved in office</i>		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)			
Test 1:	<i>0.079</i>		
Test 2:	<i>0.079</i>		
Test 3:	<i>0.079</i>		
(ATTACH TEST RECORD)			

Chad L. Harmon
Chemical Test Operator Signature

06-04-10
Date

Deb Kashner
Reviewed By

17 June 2010
Date

80-004954
P. 49 of 50
KAM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = FYPD 8164.13.00 06/09
06/04/2010 20:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:15
02 Std. Gas	0.079	20:15
03 Room Air	0.000	20:16
04 Std. Gas	0.079	20:16
05 Room Air	0.000	20:17
06 Std. Gas	0.079	20:17
07 Room Air	0.000	20:18

Lot No = 709352
Cyl No = 26
Exp Date = 12/23/2011
County = 61 Oper No. = 010606


Operator Signature
CHAD L. HARMON

Remarks:

Form 106-I8000

80-004954
p. 50 of 50
KA